



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110993

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP86818
Turbidity Calibration std, 20NTU	WP86814
Turbidity Calibration std, 40NTU	WP86813
Turbidity Calibration std, 1NTU	WP86817
Turbidity Calibration std, 5NTU	WP86816
Turbidity Calibration - CCV std, 10 NTU	WP86815

Intercept: 0.1275

Slope: 0.9804

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.254	09/16/2020	11:30
2	CAL2	1	1	1.180	09/16/2020	11:32
3	CAL3	5	1	5.120	09/16/2020	11:34
4	CAL4	10	1	9.920	09/16/2020	11:36
5	CAL5	20	1	19.200	09/16/2020	11:39
6	CAL6	40	1	39.600	09/16/2020	11:41



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB110993

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.376	09/16/2020	11:44
2	ICB		1	0.280	0.156	09/16/2020	11:46
3	CCV1	10	1	9.900	9.968	09/16/2020	11:48
4	CCB1		1	0.265	0.140	09/16/2020	11:50
5	LB110993BL		1	0.289	0.165	09/16/2020	11:52
6	L3974-36		1	4.540	4.501	09/16/2020	11:55
7	L4037-01		1	16.800	17.006	09/16/2020	11:58
8	L4037-01DUP		1	17.500	17.720	09/16/2020	12:00
9	CCV2	10	1	9.840	9.907	09/16/2020	12:02
10	CCB2		1	0.283	0.159	09/16/2020	12:05

LB110993
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-91620 WorkList ID : 142386 Department : Wet-Chemistry Date : 09-16-2020 10:58:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2020	Water	L4037-01	Turbidity	Cool 4 deg C	LIRO01	P42	UDS-6-16	09/15/2020	SM2130 B

Date/Time 9/16/2020 11:20
 Received by: Am (wet-chem)
 Relinquished by: JA (wet-chem)

Date/Time 9/16/2020 12:10
 Received by: JP Cudjoley
 Relinquished by: Am (wet-chem)