



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110995

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/16/2020	12:30
2	CAL2	1	Water	NA	NA	20.2	7.01	09/16/2020	12:31
3	CAL3	1	Water	NA	NA	20.2	10.03	09/16/2020	12:32
4	ICV	1	Water	NA	NA	20.2	7.00	09/16/2020	12:35
5	CCV1	1	Water	NA	NA	20.2	2.01	09/16/2020	12:36
6	L4030-01	1	Water	NA	NA	18.4	6.72	09/16/2020	12:39
7	L4030-01DUP	1	Water	NA	NA	18.5	6.73	09/16/2020	12:40
8	CCV2	1	Water	NA	NA	20.3	12.02	09/16/2020	12:41

87 110995

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4030 WorkList ID : 142388 Department : Wet-Chemistry Date : 09-16-2020 12:20:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/15/2020	Water	L4030-01	pH	Cool 4 deg C	APT101	P61	WGGM-20-WMV-TSF-031	09/15/2020	SM 4500-H B

Date/Time 09-16-2020 12:25
Received by: JA (Wet-Chem)
Relinquished by: JA

Date/Time 09-16-2020 13:40
Received by: JA
Relinquished by: JA flem