



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111004

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/16/2020	17:15
2	CAL2	1	Water	NA	NA	20.2	7.00	09/16/2020	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	09/16/2020	17:17
4	ICV	1	Water	NA	NA	20.2	6.98	09/16/2020	17:18
5	CCV1	1	Water	NA	NA	20.2	2.01	09/16/2020	17:19
6	L4033-05	1	Solid	20.03	20	22.8	5.88	09/16/2020	17:20
7	L4033-05DUP	1	Solid	20.04	20	22.9	5.89	09/16/2020	17:21
8	L4033-06	1	Solid	20.03	20	23.1	6.85	09/16/2020	17:22
9	L4033-07	1	Solid	20.02	20	22.8	6.87	09/16/2020	17:24
10	L4033-08	1	Solid	20.03	20	22.6	6.36	09/16/2020	17:25
11	L4033-09	1	Solid	20.04	20	22.7	6.39	09/16/2020	17:26
12	L4033-10	1	Solid	20.03	20	22.4	8.45	09/16/2020	17:27
13	L4033-11	1	Solid	20.04	20	22.4	8.30	09/16/2020	17:28
14	L4033-12	1	Solid	20.03	20	22.3	7.39	09/16/2020	17:29
15	L4033-13	1	Solid	20.04	20	22.4	7.30	09/16/2020	17:30
16	CCV2	1	Water	NA	NA	20.3	12.02	09/16/2020	17:31
17	L4033-14	1	Solid	20.03	20	22.3	7.77	09/16/2020	17:32
18	L4033-15	1	Solid	20.02	20	22.4	7.52	09/16/2020	17:35
19	L4033-16	1	Solid	20.03	20	22.6	6.91	09/16/2020	17:38
20	L4033-17	1	Solid	20.02	20	22.5	7.44	09/16/2020	17:40

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	L4033-18	1	Solid	20.03	20	22.5	7.45	09/16/2020	17:41
22	CCV3	1	Water	NA	NA	20.1	2.02	09/16/2020	17:43