


LB111016

INSTRUMENT ID: CV1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	25	0.000	25				
0.2	0.200	0.200	-0.000	1115	0.0 %	1115				
2.5	2.500	2.422	-0.078	12428	0.0 %	12428				
5.0	5.000	5.219	0.219	26663	0.0 %	26663				
7.5	7.500	7.381	-0.119	37671	0.0 %	37671				
10.0	10.000	9.993	-0.007	50966	0.0 %	50966				

LB111016 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Sep 2020 13:52:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.00007471B	PPB		17 Sep 2020 13:53:44	S	Std
0.2 - 1	30.2	1115	-	0.20007471B	PPB		17 Sep 2020 13:56:01	S	Std
2.5 - 1	52.5	12428	-	2.50007471B	PPB		17 Sep 2020 13:58:18	S	Std
5.0 - 1	55	26663	-	5.00007471B	PPB		17 Sep 2020 14:00:34	S	Std
7.5 - 1	57.5	37671	-	7.50007471B	PPB		17 Sep 2020 14:05:23	S	Std
10.0 - 1	510	50966	-	10.00007471B	PPB		17 Sep 2020 14:15:17	S	Std
ICV18 - 1	ICV18	20850	4.0766	-7471B	PPB		17 Sep 2020 14:18:01	U	SMPL
ICB18 - 1	ICB18	-144	-0.0474	-7471B	PPB		17 Sep 2020 14:20:15	U	SMPL
CCV39 - 1	CCV39	27120	5.3083	-7471B	PPB		17 Sep 2020 14:22:31	U	SMPL
CCB39 - 1	CCB39	-149	-0.0484	-7471B	PPB		17 Sep 2020 14:24:45	U	SMPL
CRA - 1	CRA	1105	0.1979	-7471B	PPB		17 Sep 2020 14:27:02	U	SMPL
HighStd - 1	HighStd	51634	10.1239	-7471B	PPB		17 Sep 2020 14:29:17	U	SMPL
ChkStd - 1	ChkStd	37646	7.3760	-7471B	PPB		17 Sep 2020 14:31:33	U	SMPL
PB131679BL - 1	PBS	-202	-0.0588	-7471B	PPB		17 Sep 2020 14:33:51	U	SMPL
PB131679BS - 1	LCSS	19664	3.8437	-7471B	PPB		17 Sep 2020 14:36:08	U	SMPL
L4032-02 - 1	SOIL-BORING	1194	0.2154	-7471B	PPB		17 Sep 2020 14:43:23	U	SMPL
L4032-02DUP - 1	SOIL-BORINGDUP	1106	0.1981	-7471B	PPB		17 Sep 2020 14:45:39	U	SMPL
L4032-02MS - 1	SOIL-BORINGMS	21078	4.1214	-7471B	PPB		17 Sep 2020 14:47:56	U	SMPL
L4032-02MSD - 1	SOIL-BORINGMSD	20985	4.1032	-7471B	PPB		17 Sep 2020 14:50:12	U	SMPL
L4033-20 - 1	51-53	9563	1.8594	-7471B	PPB		17 Sep 2020 14:52:28	U	SMPL
CCV40 - 1	CCV40	25202	4.9315	-7471B	PPB		17 Sep 2020 14:54:44	U	SMPL
CCB40 - 1	CCB40	-107	-0.0402	-7471B	PPB		17 Sep 2020 14:57:00	U	SMPL
L4033-22 - 1	54-57	15374	3.0009	-7471B	PPB		17 Sep 2020 14:59:16	U	SMPL
L4033-24 - 1	78-80	2968	0.5639	-7471B	PPB		17 Sep 2020 15:01:31	U	SMPL
L4033-26 - 1	81-84	3861	0.7393	-7471B	PPB		17 Sep 2020 15:03:48	U	SMPL
L4038-03 - 1	PAV-B36-091620-0-10	24304	4.7551	-7471B	PPB		17 Sep 2020 15:06:05	U	SMPL
L4038-07 - 1	PAV-091620-DP	18042	3.5250	-7471B	PPB		17 Sep 2020 15:08:21	U	SMPL
L4038-09 - 1	PAV-B36-091620-10-15	20114	3.9321	-7471B	PPB		17 Sep 2020 15:10:37	U	SMPL
L4032-02LX5 - 1		124	0.0052	-7471B	PPB		17 Sep 2020 15:12:54	U	SMPL
L4032-02A - 1		18895	3.6926	-7471B	PPB		17 Sep 2020 15:15:12	U	SMPL
CCV41 - 1	CCV41	26138	5.1154	-7471B	PPB		17 Sep 2020 15:17:29	U	SMPL
CCB41 - 1	CCB41	-91	-0.0370	-7471B	PPB		17 Sep 2020 15:19:45	U	SMPL