



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111031

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/17/2020	17:30
2	CAL2	1	Water	NA	NA	20.3	7.01	09/17/2020	17:31
3	CAL3	1	Water	NA	NA	20.3	10.03	09/17/2020	17:32
4	ICV	1	Water	NA	NA	20.3	7.00	09/17/2020	17:33
5	CCV1	1	Water	NA	NA	20.3	2.01	09/17/2020	17:35
6	L4053-05	1	Water	NA	NA	24.2	5.80	09/17/2020	18:00
7	L4055-01	1	Water	NA	NA	23.8	8.56	09/17/2020	18:03
8	L4055-01DUP	1	Water	NA	NA	23.9	8.57	09/17/2020	18:04
9	L4056-01	1	Water	NA	NA	24.3	7.80	09/17/2020	18:05
10	L4056-03	1	Water	NA	NA	23.6	6.98	09/17/2020	18:06
11	CCV2	1	Water	NA	NA	20.3	12.02	09/17/2020	18:10

WORKLIST(Hardcopy Internal Chain)

1811031

WorkList Name : PH L3971

WorkList ID : 142432

Department : Wet-Chemistry

Date : 09-17-2020 16:33:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2020	Water	L4053-05	pH	Cool 4 deg C	PSEG01	P51	1799-1801	09/17/2020	9040C
09/17/2020	Water	L4055-01	pH	Cool 4 deg C	PSEG01	P51	1576	09/17/2020	9040C
09/17/2020	Water	L4056-01	pH	Cool 4 deg C	PSEG01	P52	228	09/17/2020	9040C
09/17/2020	Water	L4056-03	pH	Cool 4 deg C	PSEG01	P52	229	09/17/2020	9040C

Date/Time 09-17-2020 18:11
 Received by: JN (Wet-Chem)
 Relinquished by: Peter Lumb

Date/Time 09-17-2020 18:20
 Received by: Peter Lumb
 Relinquished by: JN (Wet-Chem)