



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111036

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	09/18/2020	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	09/18/2020	11:11
3	CAL3	1	Water	NA	NA	20.2	10.03	09/18/2020	11:12
4	ICV	1	Water	NA	NA	20.2	7.02	09/18/2020	11:15
5	CCV1	1	Water	NA	NA	20.2	2.03	09/18/2020	11:16
6	L4081-01	1	Water	NA	NA	21.3	5.45	09/18/2020	11:19
7	L4081-01DUP	1	Water	NA	NA	21.4	5.46	09/18/2020	11:20
8	CCV2	1	Water	NA	NA	20.2	12.03	09/18/2020	11:22

23111036

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4081 WorkList ID : 142455 Department : Wet-Chemistry Date : 09-18-2020 10:56:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/14/2020	Water	L4081-01	pH	Cool 4 deg C	STER14	P63	JS-SYSTEM	09/14/2020	SM 4500-H B

Date/Time 09.18.2020 11:00
Received by: Jo Wet Chem
Relinquished by: Jo Wet Chem

Date/Time 09.18.2020 12:10
Received by: Jo Wet Chem
Relinquished by: Jo Wet Chem