


LB 111039

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2680e-004

C= -4.8745e-002

Rho= 0.9998476

Accept=Accepted

slope  
y-intercept

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | o/b |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.00   | 0.000  | -0.004 | -0.004 | 199   | 0.000      | 199   |       |       |       |       | -10 |
| 0.05   | 0.050  | 0.045  | -0.005 | 413   | 0.0 %      | 413   |       |       |       |       | -21 |
| 0.20   | 0.200  | 0.159  | -0.041 | 916   | 0.0 %      | 916   |       |       |       |       | 0   |
| 2.50   | 2.500  | 2.489  | -0.011 | 11190 | 0.0 %      | 11190 |       |       |       |       | 1   |
| 5.00   | 5.000  | 5.068  | 0.068  | 22562 | 0.0 %      | 22562 |       |       |       |       | 1   |
| 7.50   | 7.500  | 7.603  | 0.103  | 33738 | 0.0 %      | 33738 |       |       |       |       | 1   |
| 10.0   | 10.000 | 9.889  | -0.111 | 43818 | 0.0 %      | 43818 |       |       |       |       | -1  |

LB111039  
INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.     | Std Conc | Method    | Units | Date            | Type |
|------------|-------------|------------|-----------|----------|-----------|-------|-----------------|------|
|            | 0 S0.0      | 199        | -         |          | 0 ISM02.4 | PPB   | 9/18/2020 12:03 | Std  |
|            | 0.05 S0.05  | 413        | -         | 0.05     | ISM02.4   | PPB   | 9/18/2020 12:05 | Std  |
|            | 0.2 S0.2    | 916        | -         | 0.2      | ISM02.4   | PPB   | 9/18/2020 12:07 | Std  |
|            | 2.5 S2.5    | 11190      | -         | 2.5      | ISM02.4   | PPB   | 9/18/2020 12:10 | Std  |
|            | 5 S5.0      | 22562      | -         | 5        | ISM02.4   | PPB   | 9/18/2020 12:12 | Std  |
|            | 7.5 S7.5    | 33738      | -         | 7.5      | ISM02.4   | PPB   | 9/18/2020 12:14 | Std  |
|            | 10 S10.0    | 43818      | -         | 10       | ISM02.4   | PPB   | 9/18/2020 12:22 | Std  |
| ICV        | ICV         | 16341      | 3.6574 -  |          | ISM02.4   | PPB   | 9/18/2020 12:25 | SMPL |
| ICB        | ICB         | 164        | -0.0115 - |          | ISM02.4   | PPB   | 9/18/2020 12:30 | SMPL |
| CCV045     | CCV045      | 22131      | 4.9706 -  |          | ISM02.4   | PPB   | 9/18/2020 12:32 | SMPL |
| CCB045     | CCB045      | 106        | -0.0247 - |          | ISM02.4   | PPB   | 9/18/2020 12:34 | SMPL |
| PB131723BL | PBW723      | 140        | -0.017 -  |          | ISM02.4   | PPB   | 9/18/2020 12:39 | SMPL |
| L4039-01   | MEW0S6      | 275        | 0.0136 -  |          | ISM02.4   | PPB   | 9/18/2020 12:41 | SMPL |
| L4039-02   | MEW0S7      | -81        | -0.0671 - |          | ISM02.4   | PPB   | 9/18/2020 12:44 | SMPL |
| L4039-03   | MEW0S9      | 2          | -0.0483 - |          | ISM02.4   | PPB   | 9/18/2020 12:46 | SMPL |
| L4039-04   | MEW0S9D     | 0          | -0.0487 - |          | ISM02.4   | PPB   | 9/18/2020 12:48 | SMPL |
| L4039-05   | MEW0S9S     | 4234       | 0.9115 -  |          | ISM02.4   | PPB   | 9/18/2020 12:50 | SMPL |
| CCV046     | CCV046      | 21172      | 4.7531 -  |          | ISM02.4   | PPB   | 9/18/2020 12:53 | SMPL |
| CCB046     | CCB046      | 162        | -0.012 -  |          | ISM02.4   | PPB   | 9/18/2020 12:57 | SMPL |