

LB111047


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51982



A= 0.0000e+000

B= 1.9252e-004

C= -2.4458e-002

Rho= 0.9999905

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	72	0.000	72				
0.2	0.200	0.192	-0.008	1126	0.0 %	1126				
2.5	2.500	2.523	0.023	13235	0.0 %	13235				
5.0	5.000	4.991	-0.009	26051	0.0 %	26051				
7.5	7.500	7.521	0.021	39195	0.0 %	39195				
10.0	10.000	9.983	-0.017	51982	0.0 %	51982				

LB111047 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 18 Sep 2020 13:43:46					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	72	-	0.0000	7471B	PPB	18 Sep 2020 13:44:45	S	Std
0.2 - 1	50.2	1126	-	0.2000	7471B	PPB	18 Sep 2020 13:47:02	S	Std
2.5 - 1	52.5	13235	-	2.5000	7471B	PPB	18 Sep 2020 13:49:19	S	Std
5.0 - 1	55	26051	-	5.0000	7471B	PPB	18 Sep 2020 13:51:36	S	Std
7.5 - 1	57.5	39195	-	7.5000	7471B	PPB	18 Sep 2020 13:56:32	S	Std
10.0 - 1	510	51982	-	10.0000	7471B	PPB	18 Sep 2020 13:58:47	S	Std
ICV20 - 1	ICV20	20771	3.9743	-7471B	PPB	PPB	18 Sep 2020 14:01:26	U	SMPL
ICB20 - 1	ICB20	-115	-0.0466	-7471B	PPB	PPB	18 Sep 2020 14:03:40	U	SMPL
CCV47 - 1	CCV47	25757	4.9342	-7471B	PPB	PPB	18 Sep 2020 14:05:56	U	SMPL
CCB47 - 1	CCB47	-146	-0.0526	-7471B	PPB	PPB	18 Sep 2020 14:08:12	U	SMPL
CRA - 1	CRA	1085	0.1844	-7471B	PPB	PPB	18 Sep 2020 14:10:29	U	SMPL
HighStd - 1	HighStd	51333	9.8579	-7471B	PPB	PPB	18 Sep 2020 14:12:44	U	SMPL
ChkStd - 1	ChkStd	37680	7.2295	-7471B	PPB	PPB	18 Sep 2020 14:14:59	U	SMPL
PB131706BL - 1	PBS	-227	-0.0682	-7471B	PPB	PPB	18 Sep 2020 14:17:16	U	SMPL
PB131706BS - 1	LCSS	19667	3.7617	-7471B	PPB	PPB	18 Sep 2020 14:19:34	U	SMPL
L3870-06 - 1	SU-02-091720	7407	1.4015	-7471B	PPB	PPB	18 Sep 2020 14:21:51	U	SMPL
L3870-06MS - 1	SU-02-091720MS	24719	4.7343	-7471B	PPB	PPB	18 Sep 2020 14:26:26	U	SMPL
L3870-06MSD - 1	SU-02-091720MSD	25011	4.7905	-7471B	PPB	PPB	18 Sep 2020 14:28:43	U	SMPL
L3871-13 - 1	TR-05-091720	7073	1.3372	-7471B	PPB	PPB	18 Sep 2020 14:30:59	U	SMPL
CCV48 - 1	CCV48	25787	4.9399	-7471B	PPB	PPB	18 Sep 2020 14:33:15	U	SMPL
CCB48 - 1	CCB48	-64	-0.0368	-7471B	PPB	PPB	18 Sep 2020 14:35:31	U	SMPL
L3871-19 - 1	TR-08-091720	3411	0.6322	-7471B	PPB	PPB	18 Sep 2020 14:37:47	U	SMPL
L4057-02 - 1	JA-DRUM-COMP	240	0.0217	-7471B	PPB	PPB	18 Sep 2020 14:40:03	U	SMPL
L4059-03 - 1	PAV-B33-091720-0-10	10035	1.9074	-7471B	PPB	PPB	18 Sep 2020 14:42:19	U	SMPL
L4059-05 - 1	PAV-B33-091720-10-19	7683	1.4546	-7471B	PPB	PPB	18 Sep 2020 14:44:35	U	SMPL
L4062-01 - 1	EROC-CF-1	6320	1.1922	-7471B	PPB	PPB	18 Sep 2020 14:46:50	U	SMPL
L4062-02 - 1	EROC-CF-2	4887	0.9164	-7471B	PPB	PPB	18 Sep 2020 14:49:07	U	SMPL
L4062-03 - 1	EROC-NI4-1	196	0.0133	-7471B	PPB	PPB	18 Sep 2020 14:51:24	U	SMPL
L4062-04 - 1	EROC-NI4-2	170	0.0083	-7471B	PPB	PPB	18 Sep 2020 14:53:41	U	SMPL
L4073-01 - 1	H-1	2196	0.3983	-7471B	PPB	PPB	18 Sep 2020 14:55:58	U	SMPL
L4074-01 - 1	KTS-IB-0102-0-6	25749	4.9326	-7471B	PPB	PPB	18 Sep 2020 14:58:13	U	SMPL
CCV49 - 1	CCV49	25161	4.8194	-7471B	PPB	PPB	18 Sep 2020 15:00:29	U	SMPL
CCB49 - 1	CCB49	-91	-0.0420	-7471B	PPB	PPB	18 Sep 2020 15:02:45	U	SMPL
L4074-02 - 1	KTS-ST-0102-0-6	19334	3.6976	-7471B	PPB	PPB	18 Sep 2020 15:05:02	U	SMPL
L3870-06DUP - 1	SU-02-091720DUP	6446	1.2165	-7471B	PPB	PPB	18 Sep 2020 15:07:17	U	SMPL
L3870-06LX5 - 1		1060	0.1796	-7471B	PPB	PPB	18 Sep 2020 15:09:33	U	SMPL
L3870-06A - 1		26076	4.9956	-7471B	PPB	PPB	18 Sep 2020 15:11:49	U	SMPL
CCV50 - 1	CCV50	24744	4.7391	-7471B	PPB	PPB	18 Sep 2020 15:14:04	U	SMPL
CCB50 - 1	CCB50	-91	-0.0420	-7471B	PPB	PPB	18 Sep 2020 15:16:22	U	SMPL