


LB111061

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.6807e-004

C= 4.2647e-002

Rho= 0.9994924

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	9	0.000	9				
0.2	0.200	0.213	0.013	634	0.0 %	634				
2.5	2.500	2.496	-0.004	9152	0.0 %	9152				
5.0	5.000	5.026	0.026	18588	0.0 %	18588				
7.5	7.500	7.264	-0.236	26939	0.0 %	26939				
10.0	10.000	10.156	0.156	37728	0.0 %	37728				

LB111061 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 21 Sep 2020 08:40:40				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	9	-	0.00007470A	PPB	21 Sep 2020 08:46:32	S	Std
0.2 - 1	50.2	634	-	0.20007470A	PPB	21 Sep 2020 08:48:49	S	Std
2.5 - 1	52.5	9152	-	2.50007470A	PPB	21 Sep 2020 08:51:06	S	Std
5.0 - 1	55	18588	-	5.00007470A	PPB	21 Sep 2020 08:53:22	S	Std
7.5 - 1	57.5	26939	-	7.50007470A	PPB	21 Sep 2020 08:55:39	S	Std
10.0 - 1	510	37728	-	10.00007470A	PPB	21 Sep 2020 09:00:30	S	Std
ICV21 - 1	ICV21	13696	3.7142	-7470A	PPB	21 Sep 2020 09:03:08	U	SMPL
ICB21 - 1	ICB21	-166	-0.0019	-7470A	PPB	21 Sep 2020 09:05:23	U	SMPL
CCV51 - 1	CCV51	18847	5.0950	-7470A	PPB	21 Sep 2020 09:07:38	U	SMPL
CCB51 - 1	CCB51	-174	-0.0040	-7470A	PPB	21 Sep 2020 09:09:53	U	SMPL
CRA - 1	CRA	624	0.2099	-7470A	PPB	21 Sep 2020 09:12:08	U	SMPL
HighStd - 1	HighStd	37221	10.0206	-7470A	PPB	21 Sep 2020 09:14:23	U	SMPL
ChkStd - 1	ChkStd	24789	6.6879	-7470A	PPB	21 Sep 2020 09:16:38	U	SMPL
PB131738BL - 1	PBW	-71	0.0236	-7470A	PPB	21 Sep 2020 09:18:55	U	SMPL
PB131738BS - 1	LCSW	14291	3.8737	-7470A	PPB	21 Sep 2020 09:21:12	U	SMPL
L3870-10 - 1	SU-02-091720	51	0.0563	-7470A	PPB	21 Sep 2020 09:23:28	U	SMPL
L3870-10DUP - 1	SU-02-091720DUP	150	0.0829	-7470A	PPB	21 Sep 2020 09:25:44	U	SMPL
L3870-10MS - 1	SU-02-091720MS	14218	3.8541	-7470A	PPB	21 Sep 2020 09:28:02	U	SMPL
L3870-10MSD - 1	SU-02-091720MSD	14099	3.8222	-7470A	PPB	21 Sep 2020 09:30:18	U	SMPL
L3871-16 - 1	TR-05-091720	44	0.0544	-7470A	PPB	21 Sep 2020 09:32:34	U	SMPL
CCV52 - 1	CCV52	18758	5.0711	-7470A	PPB	21 Sep 2020 09:34:50	U	SMPL
CCB52 - 1	CCB52	-143	0.0043	-7470A	PPB	21 Sep 2020 09:37:05	U	SMPL
L3871-22 - 1	TR-08-091720	154	0.0839	-7470A	PPB	21 Sep 2020 09:39:21	U	SMPL
L4053-01 - 1	1813	432	0.1585	-7470A	PPB	21 Sep 2020 09:41:36	U	SMPL
L4053-03 - 1	VNJ-231	149	0.0826	-7470A	PPB	21 Sep 2020 09:43:51	U	SMPL
L4059-04 - 1	PAV-B33-091720-0-10	81	0.0644	-7470A	PPB	21 Sep 2020 09:46:07	U	SMPL
L4059-06 - 1	PAV-B33-091720-10-19	172	0.0888	-7470A	PPB	21 Sep 2020 09:48:23	U	SMPL
L4075-01 - 1	KTS-WC-S-IB-ST	159	0.0853	-7470A	PPB	21 Sep 2020 09:50:39	U	SMPL
PB131739BL - 1	PBW	-7	0.0408	-7470A	PPB	21 Sep 2020 10:01:35	U	SMPL
PB131739BS - 1	LCSW	12285	3.3359	-7470A	PPB	21 Sep 2020 10:06:25	U	SMPL
L4053-05 - 1	1799-1801	333	0.1319	-7470A	PPB	21 Sep 2020 10:08:41	U	SMPL
L4055-01 - 1	1576	-26	0.0357	-7470A	PPB	21 Sep 2020 10:10:57	U	SMPL
CCV53 - 1	CCV53	17265	4.6709	-7470A	PPB	21 Sep 2020 10:13:13	U	SMPL
CCB53 - 1	CCB53	-4	0.0416	-7470A	PPB	21 Sep 2020 10:15:28	U	SMPL
L4055-01DUP - 1	1576DUP	-9	0.0402	-7470A	PPB	21 Sep 2020 10:17:44	U	SMPL
L4055-01MS - 1	1576MS	12805	3.4753	-7470A	PPB	21 Sep 2020 10:19:59	U	SMPL
L4055-01MSD - 1	1576MSD	12092	3.2842	-7470A	PPB	21 Sep 2020 10:22:15	U	SMPL
L4056-01 - 1	228	14	0.0464	-7470A	PPB	21 Sep 2020 10:24:30	U	SMPL
L4056-03 - 1	229	-6	0.0410	-7470A	PPB	21 Sep 2020 10:26:46	U	SMPL
PB131738TB - 1	PB131700TB	177	0.0901	-7470A	PPB	21 Sep 2020 10:29:02	U	SMPL
PB131707TB - 1	PB131707TB	170	0.0882	-7470A	PPB	21 Sep 2020 10:31:19	U	SMPL
CCV54 - 1	CCV54	17436	4.7168	-7470A	PPB	21 Sep 2020 10:33:36	U	SMPL
CCB54 - 1	CCB54	-100	0.0158	-7470A	PPB	21 Sep 2020 10:35:53	U	SMPL
L3870-10LX5 - 1		36	0.0523	-7470A	PPB	21 Sep 2020 10:38:09	U	SMPL
L3870-10A - 1		14036	3.8053	-7470A	PPB	21 Sep 2020 10:40:24	U	SMPL
L4055-01LX5 - 1		-94	0.0174	-7470A	PPB	21 Sep 2020 10:42:40	U	SMPL
L4055-01A - 1		14252	3.8632	-7470A	PPB	21 Sep 2020 10:44:56	U	SMPL
CCV55 - 1	CCV55	17125	4.6334	-7470A	PPB	21 Sep 2020 10:47:11	U	SMPL
CCB55 - 1	CCB55	-96	0.0169	-7470A	PPB	21 Sep 2020 10:49:28	U	SMPL