



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111062

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/18/2020	17:15
2	CAL2	1	Water	NA	NA	20.3	7.00	09/18/2020	17:16
3	CAL3	1	Water	NA	NA	20.3	10.03	09/18/2020	17:18
4	ICV	1	Water	NA	NA	20.3	7.00	09/18/2020	17:20
5	CCV1	1	Water	NA	NA	20.3	2.01	09/18/2020	17:21
6	L3869-10	1	Solid	20.03	20	24.0	8.52	09/18/2020	17:22
7	L4048-02	1	Solid	20.04	20	24.1	7.40	09/18/2020	17:25
8	L4048-04	1	Solid	20.02	20	24.2	7.15	09/18/2020	17:26
9	L4048-04DUP	1	Solid	20.03	20	24.3	7.16	09/18/2020	17:29
10	CCV2	1	Water	NA	NA	20.3	12.03	09/18/2020	17:30

WORKLIST(Hardcopy Internal Chain)

28111062

WorkList Name : CORROSIONITY L3869 WorkList ID : 142480 Department : Wet-Chemistry Date : 09-18-2020 14:59:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2020	Solid	L3869-10	Corrosivity	Cool 4 deg C	PSEG05		OR-02-091820	09/18/2020	9045D
09/15/2020	Solid	L4048-02	Corrosivity	Cool 4 deg C	DIST01	P33	WC-1	09/15/2020	9045D
09/15/2020	Solid	L4048-04	Corrosivity	Cool 4 deg C	DIST01	P33	WC-2	09/15/2020	9045D

Date/Time 09-18-2020 16:25
 Received by: JP (wetchem)
 Relinquished by: JP (wetchem)

Date/Time 09-18-2020 17:30
 Received by: JP (wetchem)
 Relinquished by: JP (wetchem)
mem