



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111067

Slope : 98.0

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/21/2020	11:25
2	CAL2	1	Water	NA	NA	20.1	7.00	09/21/2020	11:26
3	CAL3	1	Water	NA	NA	20.1	10.03	09/21/2020	11:28
4	ICV	1	Water	NA	NA	20.1	7.01	09/21/2020	11:30
5	CCV1	1	Water	NA	NA	20.1	2.02	09/21/2020	11:32
6	L4094-01	1	Water	NA	NA	18.2	6.41	09/21/2020	11:35
7	L4094-01DUP	1	Water	NA	NA	18.3	6.42	09/21/2020	11:36
8	CCV2	1	Water	NA	NA	20.2	12.02	09/21/2020	11:38

13111067

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4094 WorkList ID : 142514 Department : Wet-Chemistry Date : 09-21-2020 11:13:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2020	Water	L4094-01-L	pH	Cool 4 deg C	PARS15	P53	KTS-WC-L-01	09/18/2020	9040C

Date/Time 09-21-2020 11:20
 Received by: Jg (wetchem)
 Relinquished by: KS (wetchem)

Date/Time 09-21-2020 12:23
 Received by: KS (wetchem)
 Relinquished by: Jg (wetchem)