



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111074

Ambient Barometric Pressure (mmHg): 771.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	79.9	09/21/2020	10:35
2	L4033-04		1	8	75.00	167.0	166.3	09/21/2020	11:05
3	L4033-04DUP		1	8	76.00	168.8	168.1	09/21/2020	11:40
4	L4053-01		1	9	100.00	>212.0	>212.0	09/21/2020	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB111074

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp92120 WorkList ID : 142512 Department : Wet-Chemistry Date : 09-21-2020 09:57:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/23/2020	Solid	L4033-04	Flash Point	Cool 4 deg C	PSEG01	P43	30599	09/16/2020	1010A
09/24/2020	Solid	L4053-01	Flash Point	Cool 4 deg C	PSEG01	P51	1813	09/17/2020	1010A

Date/Time 09/21/2020 10:50
Received by: KS (webster)
Relinquished by: KS (ext lab)

Date/Time 09/21/2020 12:15
Received by: KS (ext lab)
Relinquished by: KS (webster)