



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111076

Ambient Barometric Pressure (mmHg): 771.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	79.9	09/21/2020	12:45
2	L4056-01		1	8	100.00	>212.0	>212.0	09/21/2020	13:10
3	L4056-01DUP		1	8	100.00	>212.0	>212.0	09/21/2020	13:35
4	L4056-03		1	9	100.00	>212.0	>212.0	09/21/2020	13:58
5	L4094-01		1	8	100.00	>212.0	>212.0	09/21/2020	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111076

WorkList Name : FP91820 WorkList ID : 142451 Department : Wet-Chemistry Date : 09-18-2020 09:12:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/24/2020	Water	L4056-01	Flash Point	Cool 4 deg C	PSEG01	P52	228	09/17/2020	1010A
09/24/2020	Water	L4056-03	Flash Point	Cool 4 deg C	PSEG01	P52	229	09/17/2020	1010A
10/02/2020	Water	L4094-01	Flash Point	Cool 4 deg C	PARS15	P53	KTS-WC-L-01	09/18/2020	1010A

Date/Time 09/21/2020 12:20
 Received by: KS (Welchen)
 Relinquished by: Pe (ext m)

Date/Time 09/21/2020 14:25
 Received by: Pe (ext m)
 Relinquished by: KS (Welchen)