



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111079

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/21/2020	17:32
2	CAL2	1	Water	NA	NA	20.3	7.00	09/21/2020	17:33
3	CAL3	1	Water	NA	NA	20.3	10.03	09/21/2020	17:35
4	ICV	1	Water	NA	NA	20.3	7.00	09/21/2020	17:35
5	CCV1	1	Water	NA	NA	20.3	2.01	09/21/2020	17:36
6	L3864-15	1	Solid	20.03	20	20.7	8.28	09/21/2020	17:38
7	L3866-10	1	Solid	20.04	20	20.8	8.79	09/21/2020	17:39
8	L3872-10	1	Solid	20.02	20	20.6	8.40	09/21/2020	17:40
9	L4096-01	1	Solid	20.03	20	20.3	8.77	09/21/2020	17:41
10	L4096-02	1	Solid	20.04	20	20.2	9.10	09/21/2020	17:43
11	L4101-02	1	Solid	20.03	20	20.5	8.41	09/21/2020	17:45
12	L4104-04	1	Solid	20.03	20	20.8	8.70	09/21/2020	17:46
13	L4104-08	1	Solid	20.04	20	19.7	8.66	09/21/2020	17:49
14	L4104-10	1	Solid	20.03	20	20.1	8.43	09/21/2020	17:50
15	L4104-10DUP	1	Solid	20.04	20	20.3	8.44	09/21/2020	17:51
16	CCV2	1	Water	NA	NA	20.3	12.02	09/21/2020	17:53

111079

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4102 WorkList ID : 142522 Department : Wet-Chemistry Date : 09-21-2020 14:08:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/21/2020	Solid	L3864-15	Corrosivity	Cool 4 deg C	PSEG05		CL-02-092120	09/21/2020	9045D
09/21/2020	Solid	L3866-10	Corrosivity	Cool 4 deg C	PSEG05		OK-01-092120	09/21/2020	9045D
09/21/2020	Solid	L3872-10	Corrosivity	Cool 4 deg C	PSEG05		BU-02-092120	09/21/2020	9045D
09/18/2020	Solid	L4096-01	Corrosivity	Cool 4 deg C	SHAW22		B98-SOIL-IDW	09/18/2020	9045D
09/18/2020	Solid	L4096-02	Corrosivity	Cool 4 deg C	SHAW22		Z3-Z4-SOIL-IDW	09/18/2020	9045D
09/21/2020	Solid	L4101-02	Corrosivity	Cool 4 deg C	PSEG01	P63	0-6FT	09/21/2020	9045D
09/21/2020	Solid	L4104-04	Corrosivity	Cool 4 deg C	ENTE01	P63	PAV-B29-092110-0-10	09/21/2020	9045D
09/21/2020	Solid	L4104-08	Corrosivity	Cool 4 deg C	ENTE01	P63	PAV-092110-DP	09/21/2020	9045D
09/21/2020	Solid	L4104-10	Corrosivity	Cool 4 deg C	ENTE01	P63	PAV-B29-092110-10-17	09/21/2020	9045D

Date/Time 09.21.2020 16:15
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 09.21.2020 18:00
 Received by: [Signature]
 Relinquished by: [Signature]