



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111080

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/21/2020	18:00
2	CAL2	1	Water	NA	NA	20.3	7.00	09/21/2020	18:01
3	CAL3	1	Water	NA	NA	20.3	10.03	09/21/2020	18:02
4	ICV	1	Water	NA	NA	20.2	6.98	09/21/2020	18:05
5	CCV1	1	Water	NA	NA	20.3	2.02	09/21/2020	18:06
6	L4098-01	1	Solid	20.03	20	21.5	6.56	09/21/2020	18:10
7	L4098-01DUP	1	Solid	20.04	20	21.6	6.57	09/21/2020	18:11
8	CCV2	1	Water	NA	NA	20.3	12.03	09/21/2020	18:12

18111080

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L4098 WorkList ID : 142528 Department : Wet-Chemistry Date : 09-21-2020 16:43:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/21/2020	Solid	L4098-01	pH	Cool 4 deg C	PSEG01	P63	31232	09/21/2020	9045D

Date/Time 09.21.2020 16:13
Received by: Jm (wetchem)
Relinquished by: Ks (wetchem)

Date/Time 09.21.2020 18:00
Received by: Ks (wetchem)
Relinquished by: Jm (wetchem) Cym