



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111081

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/21/2020	18:00
2	CAL2	1	Water	NA	NA	20.3	7.00	09/21/2020	18:01
3	CAL3	1	Water	NA	NA	20.3	10.03	09/21/2020	18:02
4	ICV	1	Water	NA	NA	20.2	6.98	09/21/2020	18:05
5	CCV1	1	Water	NA	NA	20.3	2.02	09/21/2020	18:06
6	L4098-03	1	Water	NA	NA	20.7	5.66	09/21/2020	18:15
7	L4098-03DUP	1	Water	NA	NA	20.8	5.67	09/21/2020	18:16
8	L4098-04	1	Water	NA	NA	20.6	5.14	09/21/2020	18:19
9	L4098-05	1	Water	NA	NA	21.5	5.41	09/21/2020	18:20
10	L4098-06	1	Water	NA	NA	20.6	5.33	09/21/2020	18:21
11	L4098-07	1	Water	NA	NA	20.4	5.46	09/21/2020	18:22
12	L4098-08	1	Water	NA	NA	20.9	5.42	09/21/2020	18:25
13	L4098-09	1	Water	NA	NA	20.5	5.37	09/21/2020	18:26
14	L4098-10	1	Water	NA	NA	20.8	5.34	09/21/2020	18:29
15	L4098-11	1	Water	NA	NA	20.5	5.50	09/21/2020	18:30
16	CCV2	1	Water	NA	NA	20.2	12.03	09/21/2020	18:31
17	L4098-12	1	Water	NA	NA	20.2	5.42	09/21/2020	18:32
18	L4098-13	1	Water	NA	NA	20.5	5.27	09/21/2020	18:33
19	L4098-14	1	Water	NA	NA	20.9	5.37	09/21/2020	18:35
20	L4098-15	1	Water	NA	NA	20.7	5.43	09/21/2020	18:36

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	L4098-16	1	Water	NA	NA	20.6	8.44	09/21/2020	18:38
22	L4098-17	1	Water	NA	NA	21.5	5.35	09/21/2020	18:39
23	L4098-18	1	Water	NA	NA	21.2	4.62	09/21/2020	18:40
24	L4098-19	1	Water	NA	NA	20.1	5.61	09/21/2020	18:42
25	L4098-21	1	Water	NA	NA	20.7	8.35	09/21/2020	18:44
26	CCV3	1	Water	NA	NA	20.3	2.01	09/21/2020	18:45

VB 111081

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TCLP L4098 W WorkList ID : 142529 Department : Wet-Chemistry Date : 09-21-2020 16:44:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/21/2020	Water	L4098-03	pH	Cool 4 deg C	PSEG01	P63	31238	09/21/2020	9040C
09/21/2020	Water	L4098-04	pH	Cool 4 deg C	PSEG01	P63	31252	09/21/2020	9040C
09/21/2020	Water	L4098-05	pH	Cool 4 deg C	PSEG01	P63	31253	09/21/2020	9040C
09/21/2020	Water	L4098-06	pH	Cool 4 deg C	PSEG01	P63	31254	09/21/2020	9040C
09/21/2020	Water	L4098-07	pH	Cool 4 deg C	PSEG01	P63	31255	09/21/2020	9040C
09/21/2020	Water	L4098-08	pH	Cool 4 deg C	PSEG01	P63	31256	09/21/2020	9040C
09/21/2020	Water	L4098-09	pH	Cool 4 deg C	PSEG01	P63	31257	09/21/2020	9040C
09/21/2020	Water	L4098-10	pH	Cool 4 deg C	PSEG01	P63	31258	09/21/2020	9040C
09/21/2020	Water	L4098-11	pH	Cool 4 deg C	PSEG01	P63	31259	09/21/2020	9040C
09/21/2020	Water	L4098-12	pH	Cool 4 deg C	PSEG01	P63	31260	09/21/2020	9040C
09/21/2020	Water	L4098-13	pH	Cool 4 deg C	PSEG01	P63	31261	09/21/2020	9040C
09/21/2020	Water	L4098-14	pH	Cool 4 deg C	PSEG01	P63	31262	09/21/2020	9040C
09/21/2020	Water	L4098-15	pH	Cool 4 deg C	PSEG01	P63	31264	09/21/2020	9040C
09/21/2020	Water	L4098-16	pH	Cool 4 deg C	PSEG01	P63	31263	09/21/2020	9040C
09/21/2020	Water	L4098-17	pH	Cool 4 deg C	PSEG01	P63	31266	09/21/2020	9040C
09/21/2020	Water	L4098-18	pH	Cool 4 deg C	PSEG01	P63	31265	09/21/2020	9040C
09/21/2020	Water	L4098-19	pH	Cool 4 deg C	PSEG01	P63	DRIP-LIQUID	09/21/2020	9040C
09/21/2020	Water	L4098-21	pH	Cool 4 deg C	PSEG01	P63	SEAL-POT	09/21/2020	9040C

Date/Time 09.21.2020 16:15
 Received by: JH (wettchen)
 Relinquished by: KS (wettchen)

Date/Time 09.21.2020 18:19 C
 Received by: KS (wettchen)
 Relinquished by: JH (wettchen) *over*