


LB111088

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1528e-004

C= 1.2316e-002

Rho= 0.9999725

Accept=Accepted

slope
+1.2316e-002

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.00	0.000	0.019	0.019	31	0.000	31					
0.05	0.050					413					
0.20	0.200	0.209	0.009	913	0.0 %	913					
2.50	2.500	2.450	-0.050	11323	0.0 %	11323					
5.00	5.000	5.003	0.003	23180	0.0 %	23180					
7.50	7.500	7.536	0.036	34947	0.0 %	34947					
10.0	10.000	9.984	-0.016	46317	0.0 %	46317					

11511
2000

LB111088
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	31	-		0 ISM02.4	PPB	9/22/2020 11:50	Std
	0.2 S0.2	913	-		0.2 ISM02.4	PPB	9/22/2020 11:52	Std
	2.5 S2.5	11323	-		2.5 ISM02.4	PPB	9/22/2020 11:55	Std
	5 S5.0	23180	-		5 ISM02.4	PPB	9/22/2020 11:57	Std
	7.5 S7.5	34947	-		7.5 ISM02.4	PPB	9/22/2020 11:59	Std
	10 S10.0	46317	-		10 ISM02.4	PPB	9/22/2020 12:01	Std
ICV	ICV	17840	3.853 -		ISM02.4	PPB	9/22/2020 12:04	SMPL
ICB	ICB	-79	-0.0047 -		ISM02.4	PPB	9/22/2020 12:06	SMPL
CCV063	CCV063	23346	5.0384 -		ISM02.4	PPB	9/22/2020 12:09	SMPL
CCB063	CCB063	-85	-0.006 -		ISM02.4	PPB	9/22/2020 12:11	SMPL
PB131788BL	PBS788	46	0.0222 -		ISM02.4	PPB	9/22/2020 12:13	SMPL
L4000-01	MYAQS8	10068	2.1798 -		ISM02.4	PPB	9/22/2020 12:20	SMPL
L4000-02	MYAQS8D	9575	2.0737 -		ISM02.4	PPB	9/22/2020 12:22	SMPL
L4000-03	MYAQS8S	20709	4.4706 -		ISM02.4	PPB	9/22/2020 12:25	SMPL
L4000-04	MYAQS9	2377	0.524 -		ISM02.4	PPB	9/22/2020 12:27	SMPL
L4000-05	MYAQT0	5121	1.1148 -		ISM02.4	PPB	9/22/2020 12:29	SMPL
L4000-06	MYAQT1	1542	0.3443 -		ISM02.4	PPB	9/22/2020 12:31	SMPL
L4000-07	MYAQT2	5784	1.2575 -		ISM02.4	PPB	9/22/2020 12:34	SMPL
PB131789BL	PBS789	-18	0.0084 -		ISM02.4	PPB	9/22/2020 12:36	SMPL
L4002-01	MJLTQ0	2526	0.5561 -		ISM02.4	PPB	9/22/2020 12:38	SMPL
L4002-02	MJLTQ2	2992	0.6564 -		ISM02.4	PPB	9/22/2020 12:41	SMPL
L4002-03	MJLTQ4	2107	0.4659 -		ISM02.4	PPB	9/22/2020 12:43	SMPL
L4002-04	MJLTQ5	1434	0.321 -		ISM02.4	PPB	9/22/2020 12:45	SMPL
L4002-05	MJLTQ8	4115	0.8982 -		ISM02.4	PPB	9/22/2020 12:47	SMPL
L4002-06	MJLTQ9	2108	0.4661 -		ISM02.4	PPB	9/22/2020 12:50	SMPL
L4002-07	MJLTR0	2267	0.5004 -		ISM02.4	PPB	9/22/2020 12:52	SMPL
L4002-08	MJLTR1	2913	0.6394 -		ISM02.4	PPB	9/22/2020 12:54	SMPL
L4002-09	MJLTR2	2125	0.4698 -		ISM02.4	PPB	9/22/2020 12:56	SMPL
L4002-10	MJLTR3	2669	0.5869 -		ISM02.4	PPB	9/22/2020 12:59	SMPL
L4002-11	MJLTR4	3395	0.7432 -		ISM02.4	PPB	9/22/2020 13:01	SMPL
CCV064	CCV064	23656	5.1051 -		ISM02.4	PPB	9/22/2020 13:03	SMPL
CCB064	CCB064	-78	-0.0045 -		ISM02.4	PPB	9/22/2020 13:05	SMPL
L4002-12	MJLTR5	1836	0.4076 -		ISM02.4	PPB	9/22/2020 13:08	SMPL
L4002-13	MJLTT0	533	0.1271 -		ISM02.4	PPB	9/22/2020 13:10	SMPL
L4002-14	MJLTT1	201	0.0556 -		ISM02.4	PPB	9/22/2020 13:12	SMPL
L4002-15	MJLTT6	2306	0.5088 -		ISM02.4	PPB	9/22/2020 13:15	SMPL
L4002-16	MJLTT7	2528	0.5566 -		ISM02.4	PPB	9/22/2020 13:17	SMPL
L4002-17	MJLTY3	87	0.031 -		ISM02.4	PPB	9/22/2020 13:19	SMPL
L4002-18	MJLTY4	1782	0.396 -		ISM02.4	PPB	9/22/2020 13:21	SMPL
L4002-19	MJLTX8	1315	0.2954 -		ISM02.4	PPB	9/22/2020 13:24	SMPL
L4002-20	MJLTX8D	1233	0.2778 -		ISM02.4	PPB	9/22/2020 13:26	SMPL
L4002-21	MJLTX8S	13248	2.8644 -		ISM02.4	PPB	9/22/2020 13:28	SMPL
PB131790BL	PBS790	-18	0.0084 -		ISM02.4	PPB	9/22/2020 13:30	SMPL
L4003-01	MJLTX4	977	0.2226 -		ISM02.4	PPB	9/22/2020 13:33	SMPL
L4003-02	MJLTX5	4529	0.9873 -		ISM02.4	PPB	9/22/2020 13:35	SMPL
L4003-03	MJLTX6	2055	0.4547 -		ISM02.4	PPB	9/22/2020 13:37	SMPL

LB111088
INSTRUMENT ID : CV1

L4003-04	MJLTQ6	1767	0.3927 -	ISM02.4	PPB	9/22/2020 13:39	SMPL
L4003-05	MJLTR6	1896	0.4205 -	ISM02.4	PPB	9/22/2020 13:42	SMPL
L4003-06	MJLTR8	1710	0.3805 -	ISM02.4	PPB	9/22/2020 13:44	SMPL
L4003-07	MJLTR9	238	0.0636 -	ISM02.4	PPB	9/22/2020 13:46	SMPL
L4003-08	MJLTS0	1832	0.4067 -	ISM02.4	PPB	9/22/2020 13:48	SMPL
L4003-09	MJLTS1	2191	0.484 -	ISM02.4	PPB	9/22/2020 13:51	SMPL
CCV065	CCV065	23689	5.1122 -	ISM02.4	PPB	9/22/2020 13:53	SMPL
CCB065	CCB065	-56	0.0003 -	ISM02.4	PPB	9/22/2020 13:55	SMPL
L4003-10	MJLTS2	1359	0.3049 -	ISM02.4	PPB	9/22/2020 13:58	SMPL
L4003-11	MJLTS3	817	0.1882 -	ISM02.4	PPB	9/22/2020 14:00	SMPL
L4003-12	MJLTS6	1105	0.2502 -	ISM02.4	PPB	9/22/2020 14:02	SMPL
L4003-13	MJLTS7	3294	0.7215 -	ISM02.4	PPB	9/22/2020 14:04	SMPL
L4003-14	MJLTS8	2303	0.5081 -	ISM02.4	PPB	9/22/2020 14:07	SMPL
L4003-15	MJLTS9	1097	0.2485 -	ISM02.4	PPB	9/22/2020 14:09	SMPL
L4003-16	MJLTT2	2166	0.4786 -	ISM02.4	PPB	9/22/2020 14:11	SMPL
L4003-17	MJLTT3	1288	0.2896 -	ISM02.4	PPB	9/22/2020 14:13	SMPL
L4003-18	MJLTT8	2909	0.6386 -	ISM02.4	PPB	9/22/2020 14:16	SMPL
L4003-19	MJLW4	2840	0.6237 -	ISM02.4	PPB	9/22/2020 14:18	SMPL
L4003-20	MJLW5	2953	0.6481 -	ISM02.4	PPB	9/22/2020 14:20	SMPL
L4003-21	MJLW5D	3287	0.72 -	ISM02.4	PPB	9/22/2020 14:22	SMPL
L4003-22	MJLW5S	15378	3.323 -	ISM02.4	PPB	9/22/2020 14:25	SMPL
CCV066	CCV066	23430	5.0564 -	ISM02.4	PPB	9/22/2020 14:27	SMPL
CCB066	CCB066	-45	0.0026 -	ISM02.4	PPB	9/22/2020 14:29	SMPL