



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111093

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	09/22/2020	14:10
2	CAL2	1	Water	NA	NA	20.1	7.00	09/22/2020	14:11
3	CAL3	1	Water	NA	NA	20.1	10.03	09/22/2020	14:12
4	ICV	1	Water	NA	NA	20.1	6.99	09/22/2020	14:15
5	CCV1	1	Water	NA	NA	20.2	2.01	09/22/2020	14:16
6	L4107-02	1	Water	NA	NA	21.0	7.41	09/22/2020	14:20
7	L4107-02DUP	1	Water	NA	NA	21.1	7.42	09/22/2020	14:21
8	CCV2	1	Water	NA	NA	20.2	12.02	09/22/2020	14:22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L4107

WorkList ID : 142563

Department : Wet-Chemistry

Date : 09-22-2020 13:59:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/22/2020	Water	L4107-02 -1	Corrosivity	Cool 4 deg C	PSEG01	M11	40AA-MH	09/22/2020	9040C

Date/Time

Received by:

Relinquished by:

09/22/2020 14:00

JP / wet

JP / wet

Date/Time

Received by:

Relinquished by:

09/22/2020 15:10

JP / wet

JP / wet