



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111102

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/22/2020	11:45
2	L4098-19		1	8	100.00	>212.0	>212.0	09/22/2020	12:10
3	L4098-19DUP		1	8	100.00	>212.0	>212.0	09/22/2020	12:40
4	L4098-21		1	9	100.00	>212.0	>212.0	09/22/2020	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB111102

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp92220 WorkList ID : 142550 Department : Wet-Chemistry Date : 09-22-2020 09:05:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/05/2020	Water	L4098-19	Flash Point	Cool 4 deg C	PSEG01	P63	DRIP-LIQUID	09/21/2020	1010A
10/05/2020	Water	L4098-21	Flash Point	Cool 4 deg C	PSEG01	P63	SEAL-POT	09/21/2020	1010A

Date/Time 09/22/2020 11:15
Received by: KS
Relinquished by: JP (Wet-Chem)

Date/Time 09/22/2020 13:25
Received by: JP (Wet-Chem)
Relinquished by: KS