


LB111110

7471B

INSTRUMENT ID: CVZ



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	135	0.000	135				
0.2	0.200	0.190	-0.010	1017	0.0 %	1017				
2.5	2.500	2.380	-0.120	12850	0.0 %	12850				
5.0	5.000	5.143	0.143	27781	0.0 %	27781				
7.5	7.500	7.525	0.025	40653	0.0 %	40653				
10.0	10.000	9.936	-0.064	53685	0.0 %	53685				

LB111110 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Sep 2020 12:21:48

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	135	-	0.0000	7471B	PPB	23 Sep 2020 12:22:47	S	Std
0.2 - 1	51.2	1017	-	0.2000	7471B	PPB	23 Sep 2020 12:25:05	S	Std
2.5 - 1	52.5	12850	-	2.5000	7471B	PPB	23 Sep 2020 12:27:22	S	Std
5.0 - 1	55	27781	-	5.0000	7471B	PPB	23 Sep 2020 12:29:39	S	Std
7.5 - 1	59.5	40653	-	7.5000	7471B	PPB	23 Sep 2020 12:31:55	S	Std
10.0 - 1	61.6	53685	-	10.0000	7471B	PPB	23 Sep 2020 12:36:47	S	Std
ICV26 - 1	ICV26	21976	4.0685	-	7471B	PPB	23 Sep 2020 12:39:29	U	SMPL
ICB26 - 1	ICB26	76	0.0158	-	7471B	PPB	23 Sep 2020 12:41:44	U	SMPL
CCV75 - 1	CCV75	27842	5.1540	-	7471B	PPB	23 Sep 2020 12:44:00	U	SMPL
CCB75 - 1	CCB75	78	0.0161	-	7471B	PPB	23 Sep 2020 12:46:15	U	SMPL
CRA - 1	CRA	1012	0.1890	-	7471B	PPB	23 Sep 2020 12:48:32	U	SMPL
HighStd - 1	HighStd	53911	9.9782	-	7471B	PPB	23 Sep 2020 12:50:47	U	SMPL
ChkStd - 1	ChkStd	36769	6.8060	-	7471B	PPB	23 Sep 2020 12:53:02	U	SMPL
PB131830BL - 1	PBS	-100	-0.0168	-	7471B	PPB	23 Sep 2020 12:55:20	U	SMPL
PB131830BS - 1	LCSS	23289	4.3114	-	7471B	PPB	23 Sep 2020 12:57:37	U	SMPL
L3863-05 - 1	HD-01-092220	3159	0.5863	-	7471B	PPB	23 Sep 2020 13:04:43	U	SMPL
L3863-05DUP - 1	HD-01-092220DUP	3187	0.5915	-	7471B	PPB	23 Sep 2020 13:06:59	U	SMPL
L3863-05MS - 1	HD-01-092220MS	23621	4.3729	-	7471B	PPB	23 Sep 2020 13:09:16	U	SMPL
L3863-05MSD - 1	HD-01-092220MSD	23579	4.3651	-	7471B	PPB	23 Sep 2020 13:11:33	U	SMPL
L3870-11 - 1	SU-02-092220	16244	3.0077	-	7471B	PPB	23 Sep 2020 13:13:48	U	SMPL
CCV76 - 1	CCV76	28534	5.2820	-	7471B	PPB	23 Sep 2020 13:16:03	U	SMPL
CCB76 - 1	CCB76	143	0.0282	-	7471B	PPB	23 Sep 2020 13:18:18	U	SMPL
L4105-01 - 1	MTB-SP	4562	0.8459	-	7471B	PPB	23 Sep 2020 13:20:34	U	SMPL
L4106-01 - 1	GLOUCESTER-SP	1347	0.2510	-	7471B	PPB	23 Sep 2020 13:22:50	U	SMPL
L3863-05LX5 - 1		760	0.1423	-	7471B	PPB	23 Sep 2020 13:25:06	U	SMPL
L3863-05A - 1		24450	4.5263	-	7471B	PPB	23 Sep 2020 13:27:22	U	SMPL
CCV77 - 1	CCV77	28758	5.3235	-	7471B	PPB	23 Sep 2020 13:29:37	U	SMPL
CCB77 - 1	CCB77	134	0.0265	-	7471B	PPB	23 Sep 2020 13:31:53	U	SMPL