



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111115

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/23/2020	13:15
2	CAL2	1	Water	NA	NA	20.1	7.00	09/23/2020	13:16
3	CAL3	1	Water	NA	NA	20.2	10.03	09/23/2020	13:18
4	ICV	1	Water	NA	NA	20.1	7.00	09/23/2020	12:19
5	CCV1	1	Water	NA	NA	20.2	2.02	09/23/2020	13:20
6	L4114-01	1	Water	NA	NA	20.8	7.59	09/23/2020	13:22
7	L4114-02	1	Water	NA	NA	21.2	7.66	09/23/2020	13:24
8	L4114-03	1	Water	NA	NA	21.1	8.78	09/23/2020	13:25
9	L4114-04	1	Water	NA	NA	21.2	8.81	09/23/2020	13:26
10	L4114-05	1	Water	NA	NA	21.4	8.76	09/23/2020	13:28
11	L4114-06	1	Water	NA	NA	21.9	8.79	09/23/2020	13:30
12	L4114-07	1	Water	NA	NA	21.8	7.63	09/23/2020	13:32
13	L4114-08	1	Water	NA	NA	21.6	7.64	09/23/2020	13:33
14	L4120-05	1	Water	NA	NA	22.3	7.53	09/23/2020	13:35
15	L4120-05DUP	1	Water	NA	NA	22.4	7.54	09/23/2020	13:36
16	CCV2	1	Water	NA	NA	20.2	12.03	09/23/2020	13:39

VB 111115

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4114 WorkList ID : 142603 Department : Wet-Chemistry Date : 09-23-2020 13:03:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/23/2020	Water	L4114-01	✓ pH	Cool 4 deg C	PSEG01	M11	32667	09/23/2020	9040C
09/23/2020	Water	L4114-02	✓ pH	Cool 4 deg C	PSEG01	M11	32668	09/23/2020	9040C
09/23/2020	Water	L4114-03	✓ pH	Cool 4 deg C	PSEG01	M11	32669	09/23/2020	9040C
09/23/2020	Water	L4114-04	✓ pH	Cool 4 deg C	PSEG01	M11	32670	09/23/2020	9040C
09/23/2020	Water	L4114-05	✓ pH	Cool 4 deg C	PSEG01	M11	32671	09/23/2020	9040C
09/23/2020	Water	L4114-06	✓ pH	Cool 4 deg C	PSEG01	M11	32672	09/23/2020	9040C
09/23/2020	Water	L4114-07	✓ pH	Cool 4 deg C	PSEG01	M11	32673	09/23/2020	9040C
09/23/2020	Water	L4114-08	✓ pH	Cool 4 deg C	PSEG01	M11	32674	09/23/2020	9040C
09/23/2020	Water	L4120-05	✓ pH	Cool 4 deg C	TULL02	M11	COMPACTOR-ELECTRICAL-	09/23/2020	9040C

Date/Time 09/23/2020 16:00
 Received by: JP
 Relinquished by: JP

Date/Time 09/23/2020 13:10
 Received by: JP
 Relinquished by: JP