



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB111119

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP86965
Chlorine Calibration std, 0.1ppm	WP86963
Chlorine Calibration std, 0.2ppm	WP86964
Chlorine Calibration std, 0.8ppm	WP86966
Chlorine Calibration std, 0.0ppm	WP86962
Chlorine Calibration std, 1.6ppm	WP86967
Residual Chlorine Calibration and CCV std, 0	WP86968

Intercept: 0.0134

Slope: 1.0740

Regression: 0.999909

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	09/23/2020	15:20
2	CAL2	0.1	1	0.000	0.120	0.120	09/23/2020	15:23
3	CAL3	0.2	1	0.000	0.240	0.240	09/23/2020	15:26
4	CAL4	0.4	1	0.000	0.450	0.450	09/23/2020	15:29
5	CAL5	0.8	1	0.000	0.870	0.870	09/23/2020	15:32
6	CAL6	1.6	1	0.000	1.730	1.730	09/23/2020	15:35



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB111119

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4300	0.4300	0.3880	09/23/2020	15:38
2	ICB				1	0.0000	0.0000	0.0000	-0.0120	09/23/2020	15:41
3	CCV1			0.4	1	0.0000	0.4500	0.4500	0.4070	09/23/2020	15:44
4	CCB1				1	0.0000	0.0000	0.0000	-0.0120	09/23/2020	15:47
5	LB111119BL	50	50		1	0.0000	0.0000	0.0000	-0.0120	09/23/2020	15:50
6	LB111119BS	50	50	0.4	1	0.0000	0.4300	0.4300	0.3880	09/23/2020	15:53
7	L3974-19	50	50		1	0.0000	1.2800	1.2800	1.1790	09/23/2020	15:56
8	CCV2			0.4	1	0.0000	0.4500	0.4500	0.4070	09/23/2020	15:59
9	CCB2				1	0.0000	0.0000	0.0000	-0.0120	09/23/2020	16:02