



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111121

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/23/2020	16:35
2	CAL2	1	Water	NA	NA	20.2	7.00	09/23/2020	16:36
3	CAL3	1	Water	NA	NA	20.2	10.03	09/23/2020	16:37
4	ICV	1	Water	NA	NA	20.3	7.01	09/23/2020	16:39
5	CCV1	1	Water	NA	NA	20.2	2.02	09/23/2020	16:40
6	L3871-28	1	Solid	20.03	20	24.0	7.41	09/23/2020	16:42
7	L3871-28DUP	1	Solid	20.04	20	24.1	7.42	09/23/2020	16:43
8	L4113-01	1	Solid	20.02	20	23.8	8.12	09/23/2020	16:44
9	CCV2	1	Water	NA	NA	20.3	12.02	09/23/2020	16:45

VB111121

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L4113 WorkList ID : 142611 Department : Wet-Chemistry Date : 09-23-2020 15:14:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/23/2020	Solid	L3871-28	Corrosivity	Cool 4 deg C	PSEG05		TR-04-092320	09/23/2020	9045D
09/22/2020	Solid	L4113-01	Corrosivity	Cool 4 deg C	PARS15	M11	KTS-WC-S-0102-0-92	09/22/2020	9045D

Date/Time 09/23/2020 15:10
 Received by: JS (wet-qrm)
 Relinquished by: KS (wetchem)

Date/Time 09/23/2020 17:03
 Received by: KS (wetchem)
 Relinquished by: JS (wetchem)