


LB111127

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	51	0.000	51				
0.2	0.200	0.213	0.013	755	0.0 %	755				
2.5	2.500	2.462	-0.038	8877	0.0 %	8877				
5.0	5.000	5.122	0.122	18477	0.0 %	18477				
7.5	7.500	7.279	-0.221	26266	0.0 %	26266				
10.0	10.000	10.106	0.106	36473	0.0 %	36473				

LB111127 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Sep 2020 09:49:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7470A	PPB	24 Sep 2020 09:53:56	S	Std
0.2 - 1	50.2	755	-	0.2000	7470A	PPB	24 Sep 2020 09:56:13	S	Std
2.5 - 1	52.5	8877	-	2.5000	7470A	PPB	24 Sep 2020 09:58:31	S	Std
5.0 - 1	59	18477	-	5.0000	7470A	PPB	24 Sep 2020 10:00:48	S	Std
7.5 - 1	57.5	26266	-	7.5000	7470A	PPB	24 Sep 2020 10:03:04	S	Std
10.0 - 1	510	36473	-	10.0000	7470A	PPB	24 Sep 2020 10:07:57	S	Std
ICV27 - 1	ICV27	14568	4.0388	-	7470A	PPB	24 Sep 2020 10:10:34	U	SMPL
ICB27 - 1	ICB27	-96	-0.0230	-	7470A	PPB	24 Sep 2020 10:12:48	U	SMPL
CCV81 - 1	CCV81	17650	4.8925	-	7470A	PPB	24 Sep 2020 10:15:04	U	SMPL
CCB81 - 1	CCB81	-85	-0.0199	-	7470A	PPB	24 Sep 2020 10:17:19	U	SMPL
CRA - 1	CRA	689	0.1945	-	7470A	PPB	24 Sep 2020 10:19:35	U	SMPL
HighStd - 1	HighStd	36104	10.0041	-	7470A	PPB	24 Sep 2020 10:21:50	U	SMPL
ChkStd - 1	ChkStd	22412	6.2116	-	7470A	PPB	24 Sep 2020 10:24:06	U	SMPL
PB131846BL - 1	PBW	-161	-0.0410	-	7470A	PPB	24 Sep 2020 10:26:24	U	SMPL
PB131846BS - 1	LCSW	13710	3.8012	-	7470A	PPB	24 Sep 2020 10:28:41	U	SMPL
L3863-08 - 1	HD-01-092220	-89	-0.0210	-	7470A	PPB	24 Sep 2020 10:30:57	U	SMPL
L3863-08DUP - 1	HD-01-092220DUP	18	0.0086	-	7470A	PPB	24 Sep 2020 10:33:14	U	SMPL
L3863-08MS - 1	HD-01-092220MS	14208	3.9391	-	7470A	PPB	24 Sep 2020 10:35:31	U	SMPL
L3863-08MSD - 1	HD-01-092220MSD	12868	3.5680	-	7470A	PPB	24 Sep 2020 10:37:48	U	SMPL
L3870-15 - 1	SU-02-092220	-58	-0.0124	-	7470A	PPB	24 Sep 2020 10:40:03	U	SMPL
CCV82 - 1	CCV82	18172	5.0371	-	7470A	PPB	24 Sep 2020 10:42:18	U	SMPL
CCB82 - 1	CCB82	-90	-0.0213	-	7470A	PPB	24 Sep 2020 10:44:33	U	SMPL
L4105-02 - 1	MTB-SP	44	0.0158	-	7470A	PPB	24 Sep 2020 10:46:49	U	SMPL
L4106-02 - 1	GLOUCESTER-SP	62	0.0208	-	7470A	PPB	24 Sep 2020 10:49:05	U	SMPL
PB131857BL - 1	PBW	53	0.0183	-	7470A	PPB	24 Sep 2020 10:51:20	U	SMPL
PB131857BS - 1	LCSW	15694	4.3507	-	7470A	PPB	24 Sep 2020 10:53:36	U	SMPL
L4107-02 - 1	40AA-MH	-88	-0.0208	-	7470A	PPB	24 Sep 2020 10:55:52	U	SMPL
L4107-02DUP - 1	40AA-MHDUP	16	0.0081	-	7470A	PPB	24 Sep 2020 10:58:08	U	SMPL
L4107-02MS - 1	40AA-MHMS	13655	3.7859	-	7470A	PPB	24 Sep 2020 11:00:24	U	SMPL
L4107-02MSD - 1	40AA-MHMSD	13300	3.6876	-	7470A	PPB	24 Sep 2020 11:02:41	U	SMPL
PB131858BL - 1	PBW	-66	-0.0147	-	7470A	PPB	24 Sep 2020 11:04:59	U	SMPL
PB131858BS - 1	LCSW	16251	4.5050	-	7470A	PPB	24 Sep 2020 11:07:15	U	SMPL
CCV83 - 1	CCV83	18313	5.0762	-	7470A	PPB	24 Sep 2020 11:09:30	U	SMPL
CCB83 - 1	CCB83	-63	-0.0138	-	7470A	PPB	24 Sep 2020 11:11:46	U	SMPL
L4114-09 - 1	67-68-73-74	-89	-0.0210	-	7470A	PPB	24 Sep 2020 11:14:02	U	SMPL
L4114-09DUP - 1	67-68-73-74DUP	-57	-0.0122	-	7470A	PPB	24 Sep 2020 11:16:17	U	SMPL
L4114-09MS - 1	67-68-73-74MS	15635	4.3344	-	7470A	PPB	24 Sep 2020 11:21:10	U	SMPL
L4114-09MSD - 1	67-68-73-74MSD	15614	4.3286	-	7470A	PPB	24 Sep 2020 11:23:26	U	SMPL
PB131846TB - 1	PB131812TB	-218	-0.0568	-	7470A	PPB	24 Sep 2020 11:25:42	U	SMPL
PB131857TB - 1	PB131818TB	-36	-0.0063	-	7470A	PPB	24 Sep 2020 11:28:00	U	SMPL
PB131858TB - 1	PB131849TB	-23	-0.0027	-	7470A	PPB	24 Sep 2020 11:30:16	U	SMPL
CCV84 - 1	CCV84	18508	5.1302	-	7470A	PPB	24 Sep 2020 11:54:14	U	SMPL
CCB84 - 1	CCB84	-205	-0.0532	-	7470A	PPB	24 Sep 2020 11:56:32	U	SMPL
L3863-08LX5 - 1		-27	-0.0039	-	7470A	PPB	24 Sep 2020 11:58:49	U	SMPL
L3863-08A - 1		14825	4.1100	-	7470A	PPB	24 Sep 2020 12:01:04	U	SMPL
L4107-02LX5 - 1		-123	-0.0304	-	7470A	PPB	24 Sep 2020 12:03:19	U	SMPL
L4107-02A - 1		14563	4.0375	-	7470A	PPB	24 Sep 2020 12:05:35	U	SMPL
L4114-09LX5 - 1		-103	-0.0249	-	7470A	PPB	24 Sep 2020 12:07:51	U	SMPL
L4114-09A - 1		15752	4.3668	-	7470A	PPB	24 Sep 2020 12:10:06	U	SMPL
CCV85 - 1	CCV85	17829	4.9421	-	7470A	PPB	24 Sep 2020 12:12:22	U	SMPL
CCB85 - 1	CCB85	-102	-0.0246	-	7470A	PPB	24 Sep 2020 12:14:37	U	SMPL