



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 9/28/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:50  
In Date: 09/25/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:00  
Out Date: 09/26/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB111163

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|-------------------------|---------------------------|---------|----------|
| L4150-01 | CB7H9           | 1      | 1.15           | 9.54                    | 9.42                      | 98.6    |          |
| L4150-02 | CB7H9MS         | 2      | 1.15           | 9.54                    | 9.42                      | 98.6    |          |
| L4150-03 | CB7H9MSD        | 3      | 1.15           | 9.54                    | 9.42                      | 98.6    |          |
| L4150-04 | CB7J0           | 4      | 1.16           | 9.79                    | 9.65                      | 98.4    |          |
| L4150-05 | CB7J1           | 5      | 1.12           | 9.57                    | 9.5                       | 99.2    |          |
| L4150-06 | CB7J2           | 6      | 1.18           | 9.68                    | 9.6                       | 99.1    |          |
| L4150-07 | CB7J3           | 7      | 1.18           | 9.77                    | 9.66                      | 98.7    |          |
| L4150-08 | CB7J4           | 8      | 1.12           | 9.51                    | 9.39                      | 98.6    |          |
| L4150-09 | CB7J5           | 9      | 1.14           | 9.63                    | 9.5                       | 98.5    |          |
| L4150-10 | CB7J6           | 10     | 1.16           | 9.6                     | 9.47                      | 98.5    |          |
| L4150-11 | CB7J7           | 11     | 1.18           | 9.65                    | 9.55                      | 98.8    |          |
| L4150-12 | CB7J8           | 12     | 1.16           | 9.56                    | 9.43                      | 98.5    |          |
| L4150-13 | CB7J9           | 13     | 1.18           | 9.68                    | 9.53                      | 98.2    |          |
| L4150-14 | CB7K0           | 14     | 1.16           | 9.77                    | 9.62                      | 98.3    |          |
| L4150-15 | CB7K1           | 15     | 1.18           | 9.7                     | 9.55                      | 98.2    |          |
| L4150-16 | CB7K2           | 16     | 1.18           | 9.64                    | 9.52                      | 98.6    |          |
| L4150-17 | CB7K3           | 17     | 1.16           | 9.63                    | 9.54                      | 98.9    |          |
| L4150-18 | CB7L9           | 18     | 1.15           | 9.59                    | 9.48                      | 98.7    |          |
| L4150-19 | CB7M0           | 19     | 1.14           | 9.75                    | 9.73                      | 99.8    |          |
| L4150-20 | CB7M1           | 20     | 1.13           | 9.62                    | 9.6                       | 99.8    |          |
| L4150-21 | CB7M2           | 21     | 1.16           | 9.78                    | 9.7                       | 99.1    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$