



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB111167

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP87048
Chlorine Calibration std, 0.1ppm	WP87043
Chlorine Calibration std, 0.2ppm	WP87044
Chlorine Calibration std, 0.8ppm	WP87046
Chlorine Calibration std, 0.0ppm	WP87042
Chlorine Calibration std, 1.6ppm	WP87047
Residual Chlorine Calibration and CCV std, 0	WP87045
Total Chlorine Powder Pillows	W2249

Intercept: 0.0086

Slope: 1.0673

Regression: 0.999916

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	09/25/2020	11:30
2	CAL2	0.1	1	0.000	0.120	0.120	09/25/2020	11:33
3	CAL3	0.2	1	0.000	0.230	0.230	09/25/2020	11:36
4	CAL4	0.4	1	0.000	0.440	0.440	09/25/2020	11:39
5	CAL5	0.8	1	0.000	0.850	0.850	09/25/2020	11:42
6	CAL6	1.6	1	0.000	1.720	1.720	09/25/2020	11:45



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB111167

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4500	0.4500	0.4140	09/25/2020	11:49
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	09/25/2020	11:52
3	CCV1			0.4	1	0.0000	0.4300	0.4300	0.3950	09/25/2020	11:55
4	CCB1				1	0.0000	0.0000	0.0000	-0.0080	09/25/2020	11:58
5	LB111167BL	50	50		1	0.0000	0.0000	0.0000	-0.0080	09/25/2020	12:02
6	LB111167BS	50	50	0.4	1	0.0000	0.4500	0.4500	0.4140	09/25/2020	12:05
7	L4163-01	50	50		1	0.0000	0.0000	0.0000	-0.0080	09/25/2020	12:08
8	L4163-01DUP	50	50		1	0.0000	0.0000	0.0000	-0.0080	09/25/2020	12:11
9	CCV2			0.4	1	0.0000	0.4400	0.4400	0.4040	09/25/2020	12:14
10	CCB2				1	0.0000	0.0000	0.0000	-0.0080	09/25/2020	12:17

LB111127
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Residual Chlorine-92520 WorkList ID : 142675 Department : Wet-Chemistry Date : 09-25-2020 10:27:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/21/2020	Water	L4163-01	Residual Chlorine	Cool 4 deg C	STER17	M13	JS9600	09/21/2020	SM4500 Cl G

Date/Time 9/25/2020 10:55
 Received by: AmCwest-chem
 Relinquished by: JP CWT-chem

Date/Time 9/25/2020 12:20
 Received by: JP CWT-chem
 Relinquished by: AmCwest-chem