



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111173

Slope : 97.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	09/25/2020	17:00
2	CAL2	1	Water	NA	NA	20.4	7.00	09/25/2020	17:01
3	CAL3	1	Water	NA	NA	20.4	10.03	09/25/2020	17:02
4	ICV	1	Water	NA	NA	20.5	7.01	09/25/2020	17:05
5	CCV1	1	Water	NA	NA	20.4	2.02	09/25/2020	17:06
6	L4136-04	1	Solid	20.03	20	23.0	8.42	09/25/2020	17:10
7	L4136-04DUP	1	Solid	20.04	20	23.1	8.43	09/25/2020	17:11
8	L4136-06	1	Solid	20.02	20	23.1	8.60	09/25/2020	17:12
9	CCV2	1	Water	NA	NA	20.3	12.02	09/25/2020	17:15

78111173

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L41136 WorkList ID : 142688 Department : Wet-Chemistry Date : 09-25-2020 13:20:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/24/2020	Solid	L4136-04	Corrosivity	Cool 4 deg C	ENTE01	M12	PAV-B26-092420-0-10	09/24/2020	9045D
09/24/2020	Solid	L4136-06	Corrosivity	Cool 4 deg C	ENTE01	M12	PAV-B26-092420-10-19	09/24/2020	9045D

Date/Time 09.25.2020 13:25
 Received by: 78 (w8 - GERM)
 Relinquished by: IT(SM)

Date/Time 09.25.2020 16:25
 Received by: IT(SM)
 Relinquished by: 78 (w8) GERM