



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111181

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	09/28/2020	08:20
2	CAL2	1	Water	NA	NA	21.1	7.00	09/28/2020	08:21
3	CAL3	1	Water	NA	NA	21.1	10.03	09/28/2020	08:22
4	ICV	1	Water	NA	NA	21.1	7.00	09/28/2020	08:25
5	CCV1	1	Water	NA	NA	21.1	2.01	09/28/2020	08:26
6	L4158-04	1	Water	NA	NA	25.3	9.99	09/28/2020	08:28
7	L4158-04DUP	1	Water	NA	NA	25.4	9.98	09/28/2020	08:29
8	CCV2	1	Water	NA	NA	21.1	12.02	09/28/2020	08:30

WORKLIST(Hardcopy Internal Chain)

111181

WorkList Name : ph 14158 WorkList ID : 142707 Department : Wet-Chemistry Date : 09-28-2020 08:10:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/25/2020	Water	L4158-04	pH	Cool 4 deg C	PSEG01	M13	BLUE-FRAC-TANK	09/25/2020	9040C

Date/Time 09/28/2020 8:15
 Received by: JP
 Relinquished by: JP

Date/Time 09/28/2020 9:30
 Received by: JP
 Relinquished by: JP