



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111190

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/28/2020	12:30
2	CAL2	1	Water	NA	NA	20.2	7.00	09/28/2020	12:31
3	CAL3	1	Water	NA	NA	20.2	10.03	09/28/2020	12:33
4	ICV	1	Water	NA	NA	20.2	7.02	09/28/2020	12:35
5	CCV1	1	Water	NA	NA	20.2	2.01	09/28/2020	12:36
6	L4173-02	1	Solid	20.03	20	25.4	9.44	09/28/2020	12:38
7	L4173-02DUP	1	Solid	20.04	20	25.5	9.45	09/28/2020	12:39
8	L4173-06	1	Solid	20.02	20	25.3	9.02	09/28/2020	12:42
9	L4173-10	1	Solid	20.03	20	25.3	8.72	09/28/2020	12:43
10	CCV2	1	Water	NA	NA	20.2	12.02	09/28/2020	12:45

HB111190

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4173 WorkList ID : 142733 Department : Wet-Chemistry Date : 09-28-2020 12:21:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/25/2020	Solid	L4173-02	pH	Cool 4 deg C	PSEG01	M21	TP-1-(0-6FT)	09/25/2020	9045D
09/26/2020	Solid	L4173-06	pH	Cool 4 deg C	PSEG01	M21	TP-2-(0-6FT)	09/26/2020	9045D
09/26/2020	Solid	L4173-10	pH	Cool 4 deg C	PSEG01	M21	TP-3-(0-6FT)	09/26/2020	9045D

Date/Time 09/28/2020 12:25
 Received by: JP / wet Chem
 Relinquished by:

Date/Time 09/28/2020 16:10
 Received by: JP / wet
 Relinquished by: JP / wet