


LB111192

SFAM01.0

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.2873e-004

C= 2.4415e-002

Rho= 0.9999145

Accept=Accepted

slope
y=mx+intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.042	0.042	79	0.000	79					1
0.05	0.050					413					18
0.20	0.200	0.235	0.035	922	0.0 %	922					-2
2.50	2.500	2.452	-0.048	10613	0.0 %	10613					-1
5.00	5.000	4.926	-0.074	21430	0.0 %	21430					0
7.50	7.500	7.489	-0.011	32634	0.0 %	32634					1
10.0	10.000	10.055	0.055	43853	0.0 %	43853					

LB111192

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	79	-		0 ISM02.4	PPB	9/28/2020 14:36	Std
	0.2 S01	922	-		0.2 ISM02.4	PPB	9/28/2020 14:38	Std
	2.5 S02	10613	-		2.5 ISM02.4	PPB	9/28/2020 14:41	Std
	5 S03	21430	-		5 ISM02.4	PPB	9/28/2020 14:43	Std
	7.5 S04	32634	-		7.5 ISM02.4	PPB	9/28/2020 14:48	Std
	10 S05	43853	-		10 ISM02.4	PPB	9/28/2020 14:58	Std
ICV	ICV001	16239	3.7388 -		ISM02.4	PPB	9/28/2020 15:00	SMPL
ICB	ICB001	-67	0.0091 -		ISM02.4	PPB	9/28/2020 15:03	SMPL
CCV011	CCV011	21218	4.8777 -		ISM02.4	PPB	9/28/2020 15:05	SMPL
CCB011	CCB011	-55	0.0118 -		ISM02.4	PPB	9/28/2020 15:07	SMPL
PB131948BL	PBW948	29	0.031 -		ISM02.4	PPB	9/28/2020 15:09	SMPL
L4156-07	X2P15	13233	3.0513 -		ISM02.4	PPB	9/28/2020 15:12	SMPL
L4156-08	X2P15D	13184	3.04 -		ISM02.4	PPB	9/28/2020 15:16	SMPL
L4156-09	X2P15S	17819	4.1002 -		ISM02.4	PPB	9/28/2020 15:19	SMPL
CCV012	CCV012	20933	4.8125 -		ISM02.4	PPB	9/28/2020 15:21	SMPL
CCB012	CCB012	-51	0.0127 -		ISM02.4	PPB	9/28/2020 15:23	SMPL