



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111200

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/28/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	09/28/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	09/28/2020	17:12
4	ICV	1	Water	NA	NA	20.2	7.01	09/28/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	09/28/2020	17:19
6	L3867-12	1	Solid	20.03	20	25.2	9.17	09/28/2020	17:20
7	L3867-12DUP	1	Solid	20.04	20	25.3	9.18	09/28/2020	17:22
8	L3869-15	1	Solid	20.02	20	26.1	8.46	09/28/2020	17:24
9	CCV2	1	Water	NA	NA	20.2	12.03	09/28/2020	17:25

1111200

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3689 WorkList ID : 142741 Department : Wet-Chemistry Date : 09-28-2020 14:40:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2020	Solid	L3867-12	Corrosivity	Cool 4 deg C	PSEG05		EO-01-092820	09/28/2020	9045D
09/28/2020	Solid	L3869-15	Corrosivity	Cool 4 deg C	PSEG05		OR-03-092820	09/28/2020	9045D

Date/Time 09-28-2020 16:00
 Received by: J. C. Green
 Relinquished by: J. C. Green

Date/Time 09-28-2020 14:35
 Received by: J. C. Green
 Relinquished by: J. C. Green