


LB111201

SFAM01.0

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0447e-004

C= 8.6721e-004

Rho= 0.9999129

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.00	0.000	0.016	0.016	73	0.000	73					2
0.05	0.050					413					6
0.20	0.200	0.212	0.012	1033	0.0 %	1033					1
2.50	2.500	2.525	0.025	12344	0.0 %	12344					1
5.00	5.000	4.946	-0.054	24183	0.0 %	24183					1
7.50	7.500	7.429	-0.071	36329	0.0 %	36329					1
10.0	10.000	10.072	0.072	49256	0.0 %	49256					1

LB111201
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	73	-		0 ISM02.4	PPB	9/28/2020 16:33	Std
	0.2 S01	1033	-		0.2 ISM02.4	PPB	9/28/2020 16:35	Std
	2.5 S02	12344	-		2.5 ISM02.4	PPB	9/28/2020 16:37	Std
	5 S03	24183	-		5 ISM02.4	PPB	9/28/2020 16:39	Std
	7.5 S04	36329	-		7.5 ISM02.4	PPB	9/28/2020 16:42	Std
	10 S05	49256	-		10 ISM02.4	PPB	9/28/2020 16:47	Std
ICV	ICV002	18400	3.7632	-	ISM02.4	PPB	9/28/2020 16:49	SMPL
ICB	ICB002	-21	-0.0034	-	ISM02.4	PPB	9/28/2020 16:52	SMPL
CCV013	CCV013	24280	4.9655	-	ISM02.4	PPB	9/28/2020 16:54	SMPL
CCB013	CCB013	-31	-0.0055	-	ISM02.4	PPB	9/28/2020 16:56	SMPL
PB131949BL	PBS949	38	0.0086	-	ISM02.4	PPB	9/28/2020 16:58	SMPL
L4156-19	X2P19	22415	4.5841	-	ISM02.4	PPB	9/28/2020 17:03	SMPL
L4156-20	X2P19D	21714	4.4408	-	ISM02.4	PPB	9/28/2020 17:05	SMPL
L4156-21	X2P19S	34398	7.0343	-	ISM02.4	PPB	9/28/2020 17:12	SMPL
CCV014	CCV014	23531	4.8123	-	ISM02.4	PPB	9/28/2020 17:14	SMPL
CCB014	CCB014	-45	-0.0083	-	ISM02.4	PPB	9/28/2020 17:17	SMPL