



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111208

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	3.99	09/29/2020	11:45
2	CAL2	1	Water	NA	NA	21.1	7.00	09/29/2020	11:46
3	CAL3	1	Water	NA	NA	21.1	10.03	09/29/2020	11:47
4	ICV	1	Water	NA	NA	21.1	7.01	09/29/2020	11:48
5	CCV1	1	Water	NA	NA	21.1	2.02	09/29/2020	11:49
6	L4187-01	1	Water	NA	NA	24.2	5.25	09/29/2020	11:51
7	L4187-01DUP	1	Water	NA	NA	24.3	5.26	09/29/2020	11:52
8	CCV2	1	Water	NA	NA	20.2	12.03	09/29/2020	12:00

WORKLIST(Hardcopy Internal Chain)

VB11208

WorkList Name : PH L4187

WorkList ID : 142770

Department : Wet-Chemistry

Date : 09-29-2020 11:37:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2020	Water	L4187-01	pH	Cool 4 deg C	STER07	M13	CELLS-B/C	09/28/2020	SM 4500-H B

Date/Time 09.29.2020 11:40
 Received by: J2 Fuel Chem
 Relinquished by: J2 Fuel Chem

Date/Time 09.29.2020 12:45
 Received by: J2 Fuel Chem
 Relinquished by: J2 Fuel Chem