



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111225

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	09/30/2020	08:30
2	CAL2	1	Water	NA	NA	21.1	7.00	09/30/2020	08:31
3	CAL3	1	Water	NA	NA	21.1	10.03	09/30/2020	08:32
4	ICV	1	Water	NA	NA	21.2	7.00	09/30/2020	08:33
5	CCV1	1	Water	NA	NA	21.2	2.01	09/30/2020	08:35
6	L4189-04	1	Water	NA	NA	23.7	6.53	09/30/2020	09:16
7	L4189-04DUP	1	Water	NA	NA	23.8	6.54	09/30/2020	09:19
8	CCV2	1	Water	NA	NA	21.3	12.02	09/30/2020	09:20

VB 111225

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L4189 WorkList ID : 142777 Department : Wet-Chemistry Date : 09-29-2020 14:09:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/29/2020	Water	L4189-04	Corrosivity	Cool 4 deg C	PSEG01	M21	TP-C-2334-GW	09/29/2020	9040C

Date/Time 09/30/2020 8:10
 Received by: JP (Wet-Chem)
 Relinquished by: KS (Wet-Chem)

Date/Time 09/30/2020 11:40
 Received by: KS (Wet-Chem)
 Relinquished by: JP (Wet-Chem)