



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111226

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	09/30/2020	08:30
2	CAL2	1	Water	NA	NA	21.1	7.00	09/30/2020	08:31
3	CAL3	1	Water	NA	NA	21.1	10.03	09/30/2020	08:32
4	ICV	1	Water	NA	NA	21.2	7.00	09/30/2020	08:33
5	CCV1	1	Water	NA	NA	21.2	2.02	09/30/2020	08:35
6	L4176-01	1	Solid	20.03	20	24.6	5.72	09/30/2020	09:24
7	L4176-01DUP	1	Solid	20.04	20	24.7	5.73	09/30/2020	09:25
8	L4181-01	1	Solid	20.02	20	22.6	8.20	09/30/2020	09:26
9	L4183-01	1	Solid	20.02	20	23.4	8.09	09/30/2020	09:28
10	CCV2	1	Water	NA	NA	21.2	12.02	09/30/2020	09:29

111226

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1486 S WorkList ID : 142778 Department : Wet-Chemistry Date : 09-29-2020 14:09:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2020	Solid	L4176-01	pH	Cool 4 deg C	ENTA05	M13	AC-4	09/28/2020	9045D
09/29/2020	Solid	L4181-01	pH	Cool 4 deg C	PSEG01	M21	20744	09/29/2020	9045D
09/29/2020	Solid	L4183-01	pH	Cool 4 deg C	PSEG01	M21	32676	09/29/2020	9045D

Date/Time 09/29/2020 8:10
 Received by: JS (wetchem)
 Relinquished by: JS (wetchem)

Date/Time 09/29/2020 11:40
 Received by: JS (wetchem)
 Relinquished by: JS (wetchem)