



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111227

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	09/30/2020	08:30
2	CAL2	1	Water	NA	NA	21.1	7.00	09/30/2020	08:31
3	CAL3	1	Water	NA	NA	21.1	10.03	09/30/2020	08:32
4	ICV	1	Water	NA	NA	21.2	7.00	09/30/2020	08:33
5	CCV1	1	Water	NA	NA	21.2	2.02	09/30/2020	08:35
6	L4189-02	1	Solid	20.02	20	24.8	7.65	09/30/2020	09:36
7	L4189-02DUP	1	Solid	20.03	20	24.9	7.66	09/30/2020	08:37
8	L4190-04	1	Solid	20.02	20	22.9	8.35	09/30/2020	09:39
9	L4190-06	1	Solid	20.03	20	22.7	8.72	09/30/2020	09:41
10	L4190-12	1	Solid	20.04	20	22.8	8.89	09/30/2020	09:43
11	L4190-14	1	Solid	20.05	20	22.8	8.57	09/30/2020	09:45
12	L4190-18	1	Solid	20.02	20	22.9	8.84	09/30/2020	09:48
13	CCV2	1	Water	NA	NA	21.2	12.02	09/30/2020	09:50

28111228

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L4190 WorkList ID : 142779 Department : Wet-Chemistry Date : 09-29-2020 14:11:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/29/2020	Solid	L4189-02	Corrosivity	Cool 4 deg C	PSEG01	M21	TP-C-2334-C	09/29/2020	9045D
09/29/2020	Solid	L4190-04	Corrosivity	Cool 4 deg C	ENTE01	M13	PAV-B14-092920-0-10	09/29/2020	9045D
09/29/2020	Solid	L4190-06	Corrosivity	Cool 4 deg C	ENTE01	M13	PAV-B14-092920-10-16	09/29/2020	9045D
09/29/2020	Solid	L4190-12	Corrosivity	Cool 4 deg C	ENTE01	M13	PAV-B18-092920-0-10	09/29/2020	9045D
09/29/2020	Solid	L4190-14	Corrosivity	Cool 4 deg C	ENTE01	M13	PAV-B18-092920-10-20	09/29/2020	9045D
09/29/2020	Solid	L4190-18	Corrosivity	Cool 4 deg C	ENTE01		PAV-B18-092920-DP	09/29/2020	9045D

Date/Time 09-30-2020 8:10
 Received by: JO (Wet Chem)
 Relinquished by: KS (Wet Chem)

Date/Time 09-30-2020 11:55
 Received by: KS (Wet Chem)
 Relinquished by: JO (Wet Chem)