



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111232

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/30/2020	13:18
2	CAL2	1	Water	NA	NA	20.1	7.00	09/30/2020	13:19
3	CAL3	1	Water	NA	NA	20.1	10.03	09/30/2020	13:20
4	ICV	1	Water	NA	NA	20.2	7.00	09/30/2020	13:22
5	CCV1	1	Water	NA	NA	20.2	2.02	09/30/2020	13:25
6	L4200-02	1	Water	NA	NA	17.3	7.07	09/30/2020	13:26
7	L4200-02DUP	1	Water	NA	NA	17.4	7.08	09/30/2020	13:27
8	CCV2	1	Water	NA	NA	20.2	12.02	09/30/2020	13:30

VB 111232

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4200 W WorkList ID : 142810 Department : Wet-Chemistry Date : 09-30-2020 13:14:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/29/2020	Water	L4200-02	Corrosivity	Cool 4 deg C	ARCA05	O12	PB-2020-WW-000-01	09/29/2020	9040C

Date/Time 09.30.2020 13:15
Received by: JB West-Overmyer
Relinquished by: CP 5m1

Date/Time 09.30.2020 14:25
Received by: CP 5m1
Relinquished by: JB West-Overmyer