



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/1/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 09/30/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/01/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111233

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L4184-01	COAJ6	1	1.15	9.67	8.27	83.6	
L4184-02	COAJ7	2	1.12	9.7	8.95	91.3	
L4184-03	COAK0	3	1.13	9.52	8.24	84.7	
L4184-04	COAK6	4	1.15	9.95	8.43	82.7	
L4184-05	COAL2	5	1.18	9.6	8.02	81.2	
L4184-06	COAM1	6	1.13	9.86	8.24	81.4	
L4184-07	COAM2	7	1.14	9.86	8.9	89.0	
L4184-08	COAM5	8	1.16	9.97	9.09	90.0	
L4184-09	COAN1	9	1.1	9.93	8.54	84.3	
L4184-10	COAN7	10	1.15	9.51	8.3	85.5	
L4184-11	COAL5	11	1.16	9.97	8.74	86.0	
L4184-12	COAL6	12	1.13	9.63	7.69	77.2	
L4184-13	COAL7	13	1.19	9.98	8.39	81.9	
L4184-14	COAP0	14	1.13	9.86	8.76	87.4	
L4184-15	COAP1	15	1.13	9.6	7.64	76.9	
L4184-16	COAP2	16	1.17	9.66	8.5	86.3	
L4184-17	COAQ1	17	1.1	9.79	4.65	40.9	
L4184-18	COAS2	18	1.15	9.56	8.54	87.9	
L4184-19	VHBLK01	19	1.00	2.00	2.00	100.0	VHBLK
L4184-20	COAP4	20	1.15	9.96	7.53	72.4	
L4184-21	COAP4MS	21	1.15	9.96	7.53	72.4	
L4184-22	COAP4MSD	22	1.15	9.96	7.53	72.4	
L4184-23	COAQ5	23	1.19	9.98	2.26	12.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$