


LB111237

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0414e-004

C= 1.0736e-002

Rho= 0.9999052

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.022	0.022	55	0.000	55					1
0.05	0.050					413					1
0.20	0.200	0.213	0.013	992	0.0 %	992					7
2.50	2.500	2.506	0.006	12221	0.0 %	12221					0
5.00	5.000	4.969	-0.031	24288	0.0 %	24288					1
7.50	7.500	7.412	-0.088	36257	0.0 %	36257					1
10.0	10.000	10.078	0.078	49317	0.0 %	49317					1

LB111237
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	55	-		0 ISM02.4	PPB	9/30/2020 15:07	Std
	0.2 S0.2	992	-		0.2 ISM02.4	PPB	9/30/2020 15:09	Std
	2.5 S2.5	12221	-		2.5 ISM02.4	PPB	9/30/2020 15:11	Std
	5 S5.0	24288	-		5 ISM02.4	PPB	9/30/2020 15:14	Std
	7.5 S7.5	36257	-		7.5 ISM02.4	PPB	9/30/2020 15:16	Std
	10 S10.0	49317	-		10 ISM02.4	PPB	9/30/2020 15:18	Std
ICV	ICV	18342	3.755 -		ISM02.4	PPB	9/30/2020 15:21	SMPL
ICB	ICB	-46	0.0013 -		ISM02.4	PPB	9/30/2020 15:23	SMPL
CCV030	CCV030	24360	4.9836 -		ISM02.4	PPB	9/30/2020 15:25	SMPL
CCB030	CCB030	-49	0.0007 -		ISM02.4	PPB	9/30/2020 15:28	SMPL
PB132001BL	PBS001	19	0.0146 -		ISM02.4	PPB	9/30/2020 15:30	SMPL
L4095-01	MJLTY5	3068	0.637 -		ISM02.4	PPB	9/30/2020 15:32	SMPL
L4095-02	MJLTY6	2839	0.5903 -		ISM02.4	PPB	9/30/2020 15:34	SMPL
L4095-03	MJLTY6D	2474	0.5158 -		ISM02.4	PPB	9/30/2020 15:37	SMPL
L4095-04	MJLTY6S	14876	3.0475 -		ISM02.4	PPB	9/30/2020 15:39	SMPL
CCV031	CCV031	23791	4.8674 -		ISM02.4	PPB	9/30/2020 15:43	SMPL
CCB031	CCB031	-38	0.003 -		ISM02.4	PPB	9/30/2020 15:45	SMPL