



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111240

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.00      | 80.6      | 80.6            | 09/30/2020 | 09:45     |
| 2   | L4183-01    |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 09/30/2020 | 16:45     |
| 3   | L4183-01DUP |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 09/30/2020 | 17:10     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB 111 240

WorkList Name : fp93020

WorkList ID : 142803

Department : Wet-Chemistry

Date : 09-30-2020 12:02:12

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 10/06/2020 | Solid  | L4183-01 | F    | Cool 4 deg C | PSEG01   | M21              | 32676           | 09/29/2020   | 1010A  |

Date/Time 09/30/2020 12:05  
 Received by: KS (wehchen)  
 Relinquished by: KS (wehchen)

Date/Time 09/30/2020 17:15  
 Received by: KS (wehchen)  
 Relinquished by: KS (wehchen)