



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111241

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/30/2020	09:45
2	L4180-01		1	8	100.00	>212.0	>212.0	09/30/2020	10:10
3	L4180-01DUP		1	8	100.00	>212.0	>212.0	09/30/2020	10:35
4	L4180-02		1	9	100.00	>212.0	>212.0	09/30/2020	10:55
5	L4180-03		1	8	100.00	>212.0	>212.0	09/30/2020	11:15
6	L4180-06		1	9	100.00	>212.0	>212.0	09/30/2020	11:40
7	L4181-02		1	10	100.00	>212.0	>212.0	09/30/2020	12:05
8	L4181-03		1	8	100.00	>212.0	>212.0	09/30/2020	12:30
9	L4181-04		1	9	100.00	>212.0	>212.0	09/30/2020	12:55
10	L4181-05		1	8	100.00	>212.0	>212.0	09/30/2020	13:20
11	L4181-11		1	7	100.00	>212.0	>212.0	09/30/2020	13:45
12	L4182-01		1	9	100.00	>212.0	>212.0	09/30/2020	14:10
13	L4182-03		1	8	100.00	>212.0	>212.0	09/30/2020	14:40
14	L4182-05		1	7	100.00	>212.0	>212.0	09/30/2020	15:02
15	L4183-03		1	9	100.00	>212.0	>212.0	09/30/2020	15:32
16	L4183-05		1	7	100.00	>212.0	>212.0	09/30/2020	15:58
17	L4189-04		1	8	100.00	>212.0	>212.0	09/30/2020	16:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 111241

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp93020-1 WorkList ID : 142804 Department : Wet-Chemistry Date : 09-30-2020 09:10:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/06/2020	Water	L4180-01	Flash Point	Cool 4 deg C	PSEG01	O11	40995	09/29/2020	1010A
10/06/2020	Water	L4180-02	Flash Point	Cool 4 deg C	PSEG01	O11	20775	09/29/2020	1010A
10/06/2020	Water	L4180-03	Flash Point	Cool 4 deg C	PSEG01	O11	16784	09/29/2020	1010A
10/06/2020	Water	L4180-06	Flash Point	Cool 4 deg C	PSEG01	O11	16785-40996	09/29/2020	1010A
10/06/2020	Water	L4181-02	Flash Point	Cool 4 deg C	PSEG01	M21	16789	09/29/2020	1010A
10/06/2020	Water	L4181-03	Flash Point	Cool 4 deg C	PSEG01	M21	40976	09/29/2020	1010A
10/06/2020	Water	L4181-04	Flash Point	Cool 4 deg C	PSEG01	M21	40741	09/29/2020	1010A
10/06/2020	Water	L4181-05	Flash Point	Cool 4 deg C	PSEG01	M21	40371	09/29/2020	1010A
10/06/2020	Water	L4181-11	Flash Point	Cool 4 deg C	PSEG01	M21	05-09	09/29/2020	1010A
10/06/2020	Water	L4182-01	Flash Point	Cool 4 deg C	PSEG01	M21	32212	09/29/2020	1010A
10/06/2020	Water	L4182-03	Flash Point	Cool 4 deg C	PSEG01	M21	32681	09/29/2020	1010A
10/06/2020	Water	L4182-05	Flash Point	Cool 4 deg C	PSEG01	M21	32630	09/29/2020	1010A
10/06/2020	Water	L4183-03	Flash Point	Cool 4 deg C	PSEG01	M21	32625	09/29/2020	1010A
10/06/2020	Water	L4183-05	Flash Point	Cool 4 deg C	PSEG01	M21	32676	09/29/2020	1010A
10/06/2020	Water	L4189-04	Flash Point	Cool 4 deg C	PSEG01	M21	TP-C-2334-GW	09/29/2020	1010A

Date/Time 09/30/2020 16:30
 Received by: JS (Webster)
 Relinquished by: JS (Webster)

Date/Time 09/30/2020 09:15
 Received by: JS (Webster)
 Relinquished by: JS (Webster)