

LB 111247


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36381



A= 0.0000e+000

B= 2.7164e-004

C= -1.9764e-003

Rho= 0.9997708

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	8	0.000	8				
0.2	0.200	0.169	-0.031	629	0.0 %	629				
2.5	2.500	2.481	-0.019	9142	0.0 %	9142				
5.0	5.000	5.023	0.023	18499	0.0 %	18499				
7.5	7.500	7.646	0.146	28155	0.0 %	28155				
10.0	10.000	9.881	-0.119	36381	0.0 %	36381				

LB111247 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 01 Oct 2020 09:11:37					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	7470A	PPB	01 Oct 2020 09:15:13	S	Std
0.2 - 1	50.2	629	-	0.2000	7470A	PPB	01 Oct 2020 09:17:29	S	Std
2.5 - 1	52.5	9142	-	2.5000	7470A	PPB	01 Oct 2020 09:19:46	S	Std
5.0 - 1	55	18499	-	5.0000	7470A	PPB	01 Oct 2020 09:22:03	S	Std
7.5 - 1	57.5	28155	-	7.5000	7470A	PPB	01 Oct 2020 09:24:20	S	Std
10.0 - 1	510	36381	-	10.0000	7470A	PPB	01 Oct 2020 09:26:36	S	Std
ICV37 - 1	ICV37	13540	3.6760		7470A	PPB	01 Oct 2020 09:29:19	U	SMPL
ICB37 - 1	ICB37	-145	-0.0414		7470A	PPB	01 Oct 2020 09:31:34	U	SMPL
CCV32 - 1	CCV32	18798	5.1043		7470A	PPB	01 Oct 2020 09:33:49	U	SMPL
CCB32 - 1	CCB32	-133	-0.0381		7470A	PPB	01 Oct 2020 09:36:04	U	SMPL
CRA - 1	CRA	588	0.1577		7470A	PPB	01 Oct 2020 09:38:18	U	SMPL
HighStd - 1	HighStd	36170	9.8232		7470A	PPB	01 Oct 2020 09:40:34	U	SMPL
ChkStd - 1	ChkStd	25385	6.8936		7470A	PPB	01 Oct 2020 09:42:49	U	SMPL
PB132018BL - 1	PBW	-184	-0.0520		7470A	PPB	01 Oct 2020 09:45:05	U	SMPL
PB132018BS - 1	LCSW	17062	4.6327		7470A	PPB	01 Oct 2020 09:47:22	U	SMPL
L4180-02 - 1	20775	373	0.0993		7470A	PPB	01 Oct 2020 09:54:32	U	SMPL
L4180-02DUP - 1	20775DUP	282	0.0746		7470A	PPB	01 Oct 2020 09:56:48	U	SMPL
L4180-02MS - 1	20775MS	13717	3.7241		7470A	PPB	01 Oct 2020 10:01:36	U	SMPL
L4180-02MSD - 1	20775MSD	13412	3.6412		7470A	PPB	01 Oct 2020 10:03:53	U	SMPL
L4180-03 - 1	16784	322	0.0855		7470A	PPB	01 Oct 2020 10:06:08	U	SMPL
CCV33 - 1	CCV33	17524	4.7582		7470A	PPB	01 Oct 2020 10:08:24	U	SMPL
CCB33 - 1	CCB33	-120	-0.0346		7470A	PPB	01 Oct 2020 10:10:39	U	SMPL
L4180-06 - 1	16785-40996	241	0.0635		7470A	PPB	01 Oct 2020 10:12:55	U	SMPL
L4181-02 - 1	16789	1046	0.2822		7470A	PPB	01 Oct 2020 10:15:09	U	SMPL
L4181-03 - 1	40976	187	0.0488		7470A	PPB	01 Oct 2020 10:17:25	U	SMPL
L4181-05 - 1	40371	152	0.0393		7470A	PPB	01 Oct 2020 10:19:41	U	SMPL
L4190-21 - 1	FIELD-BLANK	-25	-0.0088		7470A	PPB	01 Oct 2020 10:21:57	U	SMPL
L4180-02LX5 - 1		64	0.0154		7470A	PPB	01 Oct 2020 10:24:13	U	SMPL
L4180-02A - 1		12888	3.4989		7470A	PPB	01 Oct 2020 10:26:30	U	SMPL
CCV34 - 1	CCV34	17930	4.8685		7470A	PPB	01 Oct 2020 10:28:46	U	SMPL
CCB34 - 1	CCB34	-121	-0.0348		7470A	PPB	01 Oct 2020 10:31:03	U	SMPL