



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111255

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP87141
Turbidity Calibration std, 20NTU	WP87137
Turbidity Calibration std, 80NTU	WP87134
Turbidity Calibration std, 40NTU	WP87136
Turbidity Calibration std, 200NTU	WP87135
Turbidity Calibration std, 1NTU	WP87140
Turbidity Calibration std, 5NTU	WP87139
Formazin turbidity 400 NTU suspension	WP87133
10 NTU Standard 500 ml	W2489

Intercept: 0.3696

Slope: 0.9517

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.271	10/01/2020	12:30
2	CAL2	1	1	1.200	10/01/2020	12:32
3	CAL3	5	1	5.180	10/01/2020	12:34
4	CAL4	10	1	9.980	10/01/2020	12:36
5	CAL5	20	1	19.200	10/01/2020	12:38
6	CAL6	40	1	38.900	10/01/2020	12:40
7	CAL7	80	1	77.900	10/01/2020	12:42
8	CAL8	200	1	188.200	10/01/2020	12:44
9	CAL9	400	1	382.000	10/01/2020	12:46



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB111255

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	10.224	10/01/2020	12:49
2	ICB		1	0.282	-0.092	10/01/2020	12:51
3	CCV1	10	1	9.820	9.930	10/01/2020	12:53
4	CCB1		1	0.277	-0.097	10/01/2020	12:55
5	LB111255BL		1	0.292	-0.082	10/01/2020	12:57
6	L3860-01		2	291.000	305.380	10/01/2020	12:59
7	L3862-01		1	62.100	64.863	10/01/2020	13:01
8	L3862-02		1	5.100	4.970	10/01/2020	13:03
9	L3862-02DUP		1	5.330	5.212	10/01/2020	13:04
10	CCV2	10	1	9.600	9.699	10/01/2020	13:06
11	CCB2		1	0.262	-0.113	10/01/2020	13:07

WORKLIST(Hardcopy Internal Chain)

LBW255

WorkList Name : Turbidity-100120 WorkList ID : 142851 Department : Wet-Chemistry Date : 10-01-2020 11:42:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2020	Water	L3860-01	Turbidity	Cool 4 deg C	LIRO01	O12	UDS-2-01	09/30/2020	SM2130 B
10/01/2020	Water	L3862-01	Turbidity	Cool 4 deg C	LIRO01	O12	UDS-4-26	09/29/2020	SM2130 B
10/01/2020	Water	L3862-02	Turbidity	Cool 4 deg C	LIRO01	O12	UDS-6-16A	09/29/2020	SM2130 B

Date/Time 10/01/2020 12:00
 Received by: AM (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 10/01/2020 13:20
 Received by: KS (wet-chem)
 Relinquished by: AM (wet-chem)