


LB111264

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.068	0.068	3	0.000	3				
0.2	0.200	0.232	0.032	610	0.0 %	610				
2.5	2.500	2.362	-0.138	8543	0.0 %	8543				
5.0	5.000	4.986	-0.014	18312	0.0 %	18312				
7.5	7.500	7.553	0.053	27868	0.0 %	27868				
10.0	10.000	9.998	-0.002	36968	0.0 %	36968				

LB111264 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Oct 2020 08:53:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	7470A	PPB	02 Oct 2020 08:54:57	S	Std
0.2 - 1	50.2	610	-	0.2000	7470A	PPB	02 Oct 2020 08:57:14	S	Std
2.5 - 1	52.5	8543	-	2.5000	7470A	PPB	02 Oct 2020 08:59:31	S	Std
5.0 - 1	55	18312	-	5.0000	7470A	PPB	02 Oct 2020 09:01:48	S	Std
7.5 - 1	57.5	27868	-	7.5000	7470A	PPB	02 Oct 2020 09:04:05	S	Std
10.0 - 1	510	36968	-	10.0000	7470A	PPB	02 Oct 2020 09:06:21	S	Std
ICV39 - 1	ICV39	13420	3.6724	-7470A		PPB	02 Oct 2020 09:09:07	U	SMPL
ICB39 - 1	ICB39	-119	0.0357	-7470A		PPB	02 Oct 2020 09:11:22	U	SMPL
CCV38 - 1	CCV38	17359	4.7305	-7470A		PPB	02 Oct 2020 09:13:37	U	SMPL
CCB38 - 1	CCB38	-90	0.0435	-7470A		PPB	02 Oct 2020 09:15:51	U	SMPL
CRA - 1	CRA	603	0.2296	-7470A		PPB	02 Oct 2020 09:18:07	U	SMPL
HighStd - 1	HighStd	34899	9.4420	-7470A		PPB	02 Oct 2020 09:20:22	U	SMPL
ChkStd - 1	ChkStd	23667	6.4249	-7470A		PPB	02 Oct 2020 09:22:37	U	SMPL
PB132032BL - 1	PBW	-152	0.0268	-7470A		PPB	02 Oct 2020 09:24:54	U	SMPL
PB132032BS - 1	LCSW	12688	3.4758	-7470A		PPB	02 Oct 2020 09:27:11	U	SMPL
L3863-12 - 1	HD-02-093020	-90	0.0435	-7470A		PPB	02 Oct 2020 09:29:27	U	SMPL
L3863-12DUP - 1	HD-02-093020DUP	0	0.0677	-7470A		PPB	02 Oct 2020 09:31:44	U	SMPL
L3863-12MS - 1	HD-02-093020MS	14016	3.8325	-7470A		PPB	02 Oct 2020 09:34:00	U	SMPL
L3863-12MSD - 1	HD-02-093020MSD	13451	3.6808	-7470A		PPB	02 Oct 2020 09:42:33	U	SMPL
L3870-20 - 1	SU-03-093020	-23	0.0615	-7470A		PPB	02 Oct 2020 09:44:48	U	SMPL
CCV39 - 1	CCV39	17130	4.6690	-7470A		PPB	02 Oct 2020 09:47:03	U	SMPL
CCB39 - 1	CCB39	-87	0.0443	-7470A		PPB	02 Oct 2020 09:49:18	U	SMPL
L4190-04 - 1	PAV-B14-092920-0-10	174	0.1144	-7470A		PPB	02 Oct 2020 09:51:33	U	SMPL
L4190-06 - 1	PAV-B14-092920-10-16	198	0.1209	-7470A		PPB	02 Oct 2020 09:53:49	U	SMPL
L4190-12 - 1	PAV-B18-092920-0-10	77	0.0883	-7470A		PPB	02 Oct 2020 09:56:04	U	SMPL
L4190-14 - 1	PAV-B18-092920-10-20	-2	0.0671	-7470A		PPB	02 Oct 2020 09:58:20	U	SMPL
L4190-18 - 1	PAV-B18-092920-DP	32	0.0763	-7470A		PPB	02 Oct 2020 10:00:36	U	SMPL
L4192-01 - 1	RBS-BD001-01	56	0.0827	-7470A		PPB	02 Oct 2020 10:02:52	U	SMPL
L4194-02 - 1	CG-SOIL	37	0.0776	-7470A		PPB	02 Oct 2020 10:05:08	U	SMPL
L4200-01 - 1	PB-2020-WS-000-01	11	0.0706	-7470A		PPB	02 Oct 2020 10:07:25	U	SMPL
CCV40 - 1	CCV40	17809	4.8514	-7470A		PPB	02 Oct 2020 10:09:42	U	SMPL
CCB40 - 1	CCB40	-69	0.0491	-7470A		PPB	02 Oct 2020 10:11:57	U	SMPL
PB132032TB - 1	PB132008TB	8	0.0698	-7470A		PPB	02 Oct 2020 10:14:13	U	SMPL
L3863-12LX5 - 1		13	0.0712	-7470A		PPB	02 Oct 2020 10:16:28	U	SMPL
L3863-12A - 1		12556	3.4404	-7470A		PPB	02 Oct 2020 10:18:43	U	SMPL
CCV41 - 1	CCV41	19198	5.2245	-7470A		PPB	02 Oct 2020 10:27:37	U	SMPL
CCB41 - 1	CCB41	-91	0.0432	-7470A		PPB	02 Oct 2020 10:29:52	U	SMPL