



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111266

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/02/2020	11:15
2	CAL2	1	Water	NA	NA	20.1	7.00	10/02/2020	11:16
3	CAL3	1	Water	NA	NA	20.1	10.03	10/02/2020	11:18
4	ICV	1	Water	NA	NA	20.1	7.00	10/02/2020	11:19
5	CCV1	1	Water	NA	NA	20.1	2.02	10/02/2020	11:20
6	L4232-01	1	Water	NA	NA	21.2	6.93	10/02/2020	11:22
7	L4232-01DUP	1	Water	NA	NA	21.9	6.94	10/02/2020	11:23
8	CCV2	1	Water	NA	NA	20.2	12.02	10/02/2020	11:25

2B711266

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4232

WorkList ID : 142889

Department : Wet-Chemistry

Date : 10-02-2020 11:06:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/01/2020	Water	L4232-01	-C pH	Cool 4 deg C	PSEG01	O21	1613	10/01/2020	9040C

Date/Time 10-02-2020 11:10
Received by: JD (wet-chem)
Relinquished by: AM (wet-chem)

Date/Time 10-02-2020 12:15
Received by: AM (wet-chem)
Relinquished by: JD (wet-chem)