



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111268

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/02/2020	11:15
2	CAL2	1	Water	NA	NA	20.1	7.00	10/02/2020	11:16
3	CAL3	1	Water	NA	NA	20.1	10.03	10/02/2020	11:18
4	ICV	1	Water	NA	NA	20.1	7.00	10/02/2020	11:19
5	CCV1	1	Water	NA	NA	20.1	2.02	10/02/2020	11:20
6	L4232-02	1	Water	NA	NA	21.2	6.93	10/02/2020	11:22
7	L4232-02DUP	1	Water	NA	NA	21.9	6.94	10/02/2020	11:23
8	CCV2	1	Water	NA	NA	20.2	12.02	10/02/2020	11:25

2B 111268

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COROSIVITY L4232 WorkList ID : 142890 Department : Wet-Chemistry Date : 10-02-2020 11:06:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/01/2020	Water	L4232-02	Corrosivity	Cool 4 deg C	PSEG01	O21	1613	10/01/2020	9040C

Date/Time 10.02.2020 11:10
Received by: Jo
Relinquished by: Am Wet-Chem

Date/Time 10.02.2020 12:13
Received by: Am Wet-Chem
Relinquished by: Jo Wet-Chem *Am*