



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/5/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 14:40
In Date: 10/02/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/03/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111270

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L4201-01	C0AJ9	1	1.14	9.68	8.48	85.9	
L4201-02	C0AK9	2	1.16	9.96	8.97	88.8	
L4201-03	C0AL4	3	1.13	9.56	8.4	86.2	
L4201-04	C0AM4	4	1.13	9.89	8.2	80.7	
L4201-05	C0AN4	5	1.18	9.88	8.55	84.7	
L4201-06	C0AN9	6	1.17	9.57	8.36	85.6	
L4201-07	C0AS3	7	1.12	9.51	8.16	83.9	
L4201-08	C0AS4	8	1.15	9.95	9.26	92.2	
L4201-09	C0AL3	9	1.13	9.86	8.4	83.3	
L4201-10	C0AL8	10	1.13	9.57	7.68	77.6	
L4201-11	C0AL9	11	1.14	9.98	8.19	79.8	
L4201-12	C0AN8	12	1.13	9.77	9.06	91.8	
L4201-13	C0AP3	13	1.13	9.62	5.93	56.5	
L4201-14	C0AQ6	14	1.14	9.98	2.6	16.5	
L4201-15	VHBLK01	15	1.00	2.00	2.00	100.0	VHBLK
L4201-16	C0AJ8	16	1.13	9.82	8.14	80.7	
L4201-17	C0AK3	17	1.13	9.52	8.64	89.5	
L4201-18	C0AK8	18	1.13	9.75	8.86	89.7	
L4201-19	C0AL1	19	1.13	9.91	8.42	83.0	
L4201-20	C0AM3	20	1.16	9.55	9.19	95.7	
L4201-21	C0AM8	21	1.13	9.87	9.26	93.0	
L4201-22	C0AM8MS	22	1.13	9.87	9.26	93.0	
L4201-23	C0AM8MSD	23	1.13	9.87	9.26	93.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$