



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111274

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP87199
Turbidity Calibration std, 20NTU	WP87195
Turbidity Calibration std, 40NTU	WP87194
Turbidity Calibration std, 1NTU	WP87198
Turbidity Calibration std, 5NTU	WP87197
Turbidity Calibration - CCV std, 10 NTU	WP87196
10 NTU Standard 500 ml	W2489

Intercept: 0.272

Slope: 0.9535

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.275	10/02/2020	11:50
2	CAL2	1	1	0.982	10/02/2020	11:52
3	CAL3	5	1	5.120	10/02/2020	11:54
4	CAL4	10	1	9.720	10/02/2020	11:56
5	CAL5	20	1	19.800	10/02/2020	11:58
6	CAL6	40	1	38.200	10/02/2020	12:00



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB111274

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.200	10.412	10/02/2020	12:03
2	ICB		1	0.281	0.009	10/02/2020	12:05
3	CCV1	10	1	9.560	9.741	10/02/2020	12:07
4	CCB1		1	0.293	0.022	10/02/2020	12:09
5	LB111274BL		1	0.287	0.016	10/02/2020	12:12
6	L4252-01		1	34.100	35.478	10/02/2020	12:14
7	L4252-01DUP		1	33.200	34.534	10/02/2020	12:16
8	CCV2	10	1	9.980	10.181	10/02/2020	12:19
9	CCB2		1	0.296	0.025	10/02/2020	12:22

LB11274 WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-100220 WorkList ID : 142883 Department : Wet-Chemistry Date : 10-02-2020 10:45:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/03/2020	Water	L4252-01	Turbidity	Cool 4 deg C	LIRO01	O11	UDS-4-40	10/01/2020	SM2130 B

Date/Time 10/02/2020 11:20
 Received by: AM(wet-chem)
 Relinquished by: JD(wet-chem)

Date/Time 10/02/2020 12:40
 Received by: JP(wet-chem)
 Relinquished by: AM(wet-chem)