


LB 111277

ISM02.4

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.3103e-004

C= -6.4861e-002

Rho= 0.9999918

Accept=Accepted

slope
4.15 Intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.009	-0.009	241	0.000	241					
0.05	0.050					413					
0.20	0.200	0.187	-0.013	1092	0.0 %	1092					
2.50	2.500	2.530	0.030	11231	0.0 %	11231					
5.00	5.000	4.994	-0.006	21899	0.0 %	21899					
7.50	7.500	7.508	0.008	32779	0.0 %	32779					
10.0	10.000	9.989	-0.011	43519	0.0 %	43519					

0.00
-7
1
0
0
0

LB111277

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	241	-		0 ISM02.4	PPB	10/2/2020 14:39	Std
	0.2 S0.2	1092	-		0.2 ISM02.4	PPB	10/2/2020 14:42	Std
	2.5 S2.5	11231	-		2.5 ISM02.4	PPB	10/2/2020 14:44	Std
	5 S5.0	21899	-		5 ISM02.4	PPB	10/2/2020 14:46	Std
	7.5 S7.5	32779	-		7.5 ISM02.4	PPB	10/2/2020 14:48	Std
	10 S10.0	43519	-		10 ISM02.4	PPB	10/2/2020 14:51	Std
ICV	ICV	17068	3.8784 -		ISM02.4	PPB	10/2/2020 14:53	SMPL
ICB	ICB	406	0.0289 -		ISM02.4	PPB	10/2/2020 14:56	SMPL
CCV046	CCV046	21629	4.9321 -		ISM02.4	PPB	10/2/2020 14:58	SMPL
CCB046	CCB046	402	0.028 -		ISM02.4	PPB	10/2/2020 15:00	SMPL
PB132052BL	PBS052	2	-0.0644 -		ISM02.4	PPB	10/2/2020 15:02	SMPL
L4167-01	MCOAB4	2019	0.4016 -		ISM02.4	PPB	10/2/2020 15:05	SMPL
L4167-02	MCOAB9	3955	0.8489 -		ISM02.4	PPB	10/2/2020 15:07	SMPL
L4167-03	MCOAC2	287	0.0014 -		ISM02.4	PPB	10/2/2020 15:09	SMPL
L4167-04	MCOAC3	150	-0.0302 -		ISM02.4	PPB	10/2/2020 15:11	SMPL
L4167-05	MCOAC4	219	-0.0143 -		ISM02.4	PPB	10/2/2020 15:14	SMPL
L4167-06	MCOAC5	122	-0.0367 -		ISM02.4	PPB	10/2/2020 15:16	SMPL
L4167-07	MCOAC7	3304	0.6985 -		ISM02.4	PPB	10/2/2020 15:18	SMPL
L4167-08	MCOAC9	6699	1.4828 -		ISM02.4	PPB	10/2/2020 15:21	SMPL
L4167-09	MCOAD1	4565	0.9898 -		ISM02.4	PPB	10/2/2020 15:23	SMPL
L4167-10	MCOAE2	8821	1.9731 -		ISM02.4	PPB	10/2/2020 15:25	SMPL
L4167-11	MCOAE4	8980	2.0098 -		ISM02.4	PPB	10/2/2020 15:27	SMPL
L4167-16	MCOAE4D	8864	1.983 -		ISM02.4	PPB	10/2/2020 15:30	SMPL
PB132053BL	PBS053	15	-0.0614 -		ISM02.4	PPB	10/2/2020 15:39	SMPL
L4206-01	MBEQ37	8889	1.9888 -		ISM02.4	PPB	10/2/2020 15:41	SMPL
L4206-02	MBEQ38	9612	2.1558 -		ISM02.4	PPB	10/2/2020 15:43	SMPL
L4206-03	MBEQ39	12832	2.8997 -		ISM02.4	PPB	10/2/2020 15:46	SMPL
L4206-04	MBEQ40	10261	2.3057 -		ISM02.4	PPB	10/2/2020 15:48	SMPL
L4206-05	MBEQ41	23822	5.4387 -		ISM02.4	PPB	10/2/2020 15:50	SMPL
CCV047	CCV047	20650	4.7059 -		ISM02.4	PPB	10/2/2020 15:52	SMPL
CCB047	CCB047	388	0.0248 -		ISM02.4	PPB	10/2/2020 15:55	SMPL
L4206-06	MBEQ42	10431	2.345 -		ISM02.4	PPB	10/2/2020 15:57	SMPL
L4206-07	MBEQ43	9477	2.1246 -		ISM02.4	PPB	10/2/2020 15:59	SMPL
L4206-08	MBEQ44	12444	2.8101 -		ISM02.4	PPB	10/2/2020 16:02	SMPL
L4206-09	MBEQ45	8938	2.0001 -		ISM02.4	PPB	10/2/2020 16:04	SMPL
L4206-10	MBEQ45D	9470	2.123 -		ISM02.4	PPB	10/2/2020 16:06	SMPL
L4206-11	MBEQ45S	19231	4.3781 -		ISM02.4	PPB	10/2/2020 16:11	SMPL
L4206-12	MBEQ46	1005	0.1673 -		ISM02.4	PPB	10/2/2020 16:13	SMPL
L4206-13	MBEQ72	10147	2.2794 -		ISM02.4	PPB	10/2/2020 16:16	SMPL
L4206-14	MBEQ79	1475	0.2759 -		ISM02.4	PPB	10/2/2020 16:18	SMPL
L4206-15	MBEQA2	10969	2.4693 -		ISM02.4	PPB	10/2/2020 16:20	SMPL
L4206-16	MBEQA3	135	-0.0337 -		ISM02.4	PPB	10/2/2020 16:22	SMPL
L4206-17	MBEQA4	26881	6.1455 -		ISM02.4	PPB	10/2/2020 16:25	SMPL
L4206-18	MBEQA5	22629	5.1631 -		ISM02.4	PPB	10/2/2020 16:27	SMPL
L4206-19	MBEQA6	10769	2.4231 -		ISM02.4	PPB	10/2/2020 16:29	SMPL
L4206-20	MBEQB0	12540	2.8323 -		ISM02.4	PPB	10/2/2020 16:31	SMPL

LB111277

INSTRUMENT ID : CV1

L4206-21	MBEQB1	5938	1.307 -	ISM02.4	PPB	10/2/2020 16:34	SMPL
L4206-22	MBEQB2	43434	9.9697 -	ISM02.4	PPB	10/2/2020 16:36	SMPL
L4167-17	MC0AE4S	18311	4.1655 -	ISM02.4	PPB	10/2/2020 16:39	SMPL
CCV048	CCV048	20757	4.7306 -	ISM02.4	PPB	10/2/2020 16:41	SMPL
CCB048	CCB048	345	0.0148 -	ISM02.4	PPB	10/2/2020 16:44	SMPL