



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111280

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/02/2020	15:30
2	CAL2	1	Water	NA	NA	20.2	7.01	10/02/2020	15:31
3	CAL3	1	Water	NA	NA	20.2	10.03	10/02/2020	15:32
4	ICV	1	Water	NA	NA	20.2	7.00	10/02/2020	15:35
5	CCV1	1	Water	NA	NA	20.2	2.01	10/02/2020	15:36
6	L4219-08	1	Solid	20.03	20	21.7	9.00	10/02/2020	15:39
7	L4221-05	1	Solid	20.04	20	21.7	8.69	10/02/2020	15:40
8	L4226-04	1	Solid	20.03	20	22.1	8.49	10/02/2020	15:42
9	L4242-04	1	Solid	20.04	20	22.0	8.59	10/02/2020	15:43
10	L4242-06	1	Solid	20.02	20	22.1	8.50	10/02/2020	15:45
11	L4242-12	1	Solid	20.03	20	22.3	9.12	10/02/2020	15:48
12	L4242-14	1	Solid	20.04	20	22.1	8.62	10/02/2020	15:50
13	L4242-20	1	Solid	20.03	20	21.9	8.63	10/02/2020	16:00
14	L4247-11	1	Solid	20.04	20	22.0	8.30	10/02/2020	16:02
15	L4247-13	1	Solid	20.03	20	22.1	8.36	10/02/2020	16:05
16	CCV2	1	Water	NA	NA	20.2	12.02	10/02/2020	16:06
17	L4247-13DUP	1	Solid	20.04	20	22.2	8.37	10/02/2020	16:10
18	CCV3	1	Water	NA	NA	20.2	2.02	10/02/2020	16:11

WORKLIST(Hardcopy Internal Chain)

VB 111230

WorkList Name : CORROSIONITY L4242 WorkList ID : 142888 Department : Wet-Chemistry Date : 10-02-2020 11:05:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/01/2020	Solid	L4219-08	Corrosivity	Cool 4 deg C	PSEG05	O21	CL-01-100120	10/01/2020	9045D
10/01/2020	Solid	L4221-05	Corrosivity	Cool 4 deg C	PSEG05	O21	OK-02-100120	10/01/2020	9045D
10/01/2020	Solid	L4226-04	Corrosivity	Cool 4 deg C	PSEG05	O21	TR-05-100120	10/01/2020	9045D
10/01/2020	Solid	L4242-04	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B24-100120-0-10	10/01/2020	9045D
10/01/2020	Solid	L4242-06	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B24-100120-10-18	10/01/2020	9045D
10/01/2020	Solid	L4242-12	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B17-100120-0-10	10/01/2020	9045D
10/01/2020	Solid	L4242-14	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B17-100120-10-19	10/01/2020	9045D
10/01/2020	Solid	L4242-20	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B24-100120-DP	10/01/2020	9045D
10/01/2020	Solid	L4247-11	Corrosivity	Cool 4 deg C	URSC01	O13	B-18(0-1)	10/01/2020	9045D
10/01/2020	Solid	L4247-13	Corrosivity	Cool 4 deg C	URSC01	O13	B-19(4-5)	10/01/2020	9045D

Date/Time 10-02-2020 11:12
 Received by: JP (wetchem)
 Relinquished by: KS

Date/Time 10-02-2020 16:30
 Received by: KS
 Relinquished by: JP (wetchem)